



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S : Rainbow Diagnostics Unit of SC-45, C Block, Near Union Bank, Shalimar Garden Extn. II, Sahibabad, Ghaziabad	Certificate No.	CC2239206000042600F
		Service request No & date	A/01/11/12/2020
		Date of calibration	11/12/2020
		Valid up to	10/06/2021
		Date of issue	11/12/2020

Instrument Details

Instrument name	Deep Freezer	Sr. No.	-----
Make	-----	Location	-----
Range / Size	-23 to 7 °C	Accuracy	----
Least Count	N/A	Visual Inspection	OK
I.D. No.	DP - 01		

Detail of reference standards & Major equipments used

Equipment Name	RTD With Digital Indicator		
Make	YCT / SBI		
Model / SR No.	11080368		
Certificate No.	TL/020/181.2.1		
Calibration Validity	01/03/2021		
Calibration by	Tempens		

Environmental Condition	Temperature	25±4 °C	Calibration Reference	ITS:90
	Relative Humidity	50±20%	Calibration Procedure	CP-06T

Calibration Results

Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	-22.35	-22.0	±0.2 °C
02.	-18.14	-18.0	
03.	-10.36	-10.0	
04.	0.04	0.0	

Remarks:

- ❖ This result of calibration refer to only to the particular item submitted for calibration.
- ❖ The above result are valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By
(Calibration Engg./TM)
(D.B.SAXENA)
Form No.-32



----End of Report----

Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1