
 <p><b>JOTHI LABS</b></p>	<p><b>Jothi X-RAY, ECG &amp; Laboratory</b></p> <p><b>CORRECTIVE ACTION REPORT</b></p>	Form No.	JXEL / QAS-BC/230
		Issue No.	01
		Rev No.	0
		Issue Date	12.12.2020

CAR No.: 01	Date: 12-12-2020
-------------	------------------

Sample <input type="checkbox"/>	Process <input checked="" type="checkbox"/>	System <input type="checkbox"/>	Cust. Complaint <input type="checkbox"/>
---------------------------------	---	---------------------------------	--


**Detail of Non-Conformance:**

We used Bio system analyser for S.Bilirubin, Uric acid & S.Cholestrol. so the result value was high.

Signature:  Date: 12.12.2020


**Root Cause Analysis :**

We started using Bio chemistry Internal Quality Control sample test.

Signature:  Date: 12.12.2020


**Corrective Action :**

We have changed the brand of the Reagent kit to correct result for S.Bilirubin, Uric acid & S.Cholestrol.

Signature:  Date: 12.12.2020

**Verification & Effectiveness of Corrective Action:**

We get counseling/clarification from the manager of the company. Also , we got accurate Internal Quality Control report.

Signature:  Date: 12.12.2020