



**CMC VIROEQAS**

PC-1034

**Under National AIDS Control Organisation (NACO)**  
**PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu**  
**Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455**

Comments:

Participants are scored based on qualitative result.

**Assigned Value:**

Assigned value is determined using multiple assays/multiple testing of the same analyte. The proficiency testing material is a pooled lyophilized plasma sample prepared by following institutional proficiency testing standard operating protocol (PT-SOP) by trained staff. None of the work related to testing, preparation and packaging of sample is subcontracted. The homogeneity of the PT material is determined using multiple testing of multiple aliquots. The stability of the testing material is determined by assessing the reactivity of the specimen till the closing date.

Enquires: For queries, please contact CMCVIROEQAS coordinator at email: [nrl.cmc@gmail.com](mailto:nrl.cmc@gmail.com)

**Name of CMCVIROEQAS coordinator**

**Signature**


**Dr. Rajesh Kannangai**

**Department of Clinical Virology**

**Report Dispatch Date: 18-12-2020**

Report authorized by: CMCVIROEQAS Coordinator

**END OF REPORT**



**CMC VIROEQAS**

PC-1034

**Under National AIDS Control Organisation (NACO)**

**PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu**

**Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455**

**PANEL: HIV Serology**

Name of the laboratory: ICTC, District Hospital, Nedumangadu, Thiruvananthapuram

Distribution Round -2: Year 2019

Result receiving month: December 2019

SL NO	SAMPLE ID	INTENDED RESULT	YOUR RESULT	MATCHING YES / NO	% MATCHING
1	S021901	NEGATIVE	NEGATIVE	YES	100
2	S021902	POSITIVE FOR HIV-1	POSITIVE FOR HIV-1	YES	100
3	S021903	POSITIVE FOR HIV-1	POSITIVE FOR HIV-1	YES	100
4	S021904	NEGATIVE	NEGATIVE	YES	100

**REMARKS: PERFORMANCE IS SATISFACTORY**

PT Unit, Department of Clinical Virology, Christian Medical College, Vellore

The data in this CMCVIROEQAS reports are confidential