




ULR No. : CC29372000000801P
Certificate No : TVCS 20/11/189 - 01

MECHANICAL CALIBRATION (Volume)				
Calibration Results				
2. Middle Volume :	25	μl	No. of Measurements :	10
<input type="text" value="25.01"/>	<input type="text" value="24.93"/>		<input type="text" value="24.88"/>	<input type="text" value="24.91"/>
<input type="text" value="24.88"/>	<input type="text" value="24.89"/>		<input type="text" value="24.82"/>	<input type="text" value="24.87"/>
<input type="text" value="24.83"/>	<input type="text" value="24.96"/>			
Mean Value :	24.90	μl		
Error Limits(\pm)				
Systematic Error :	-0.10	μl	0.20	μl
Systematic Error :	-0.41	%	1.00	%
Random Error :	0.06	μl	0.10	μl
Random Error :	0.23	%	0.50	%
Measurement Uncertainty :	\pm		0.14 μl	
3. Nominal Volume :	50	μl	No. of Measurements :	10
<input type="text" value="50.08"/>	<input type="text" value="50.12"/>		<input type="text" value="50.14"/>	<input type="text" value="49.99"/>
<input type="text" value="50.11"/>	<input type="text" value="50.04"/>		<input type="text" value="50.23"/>	<input type="text" value="50.24"/>
<input type="text" value="50.21"/>	<input type="text" value="50.23"/>			
Mean Value :	50.14	μl		
Error Limits(\pm)				
Systematic Error :	0.14	μl	0.50	μl
Systematic Error :	0.28	%	1.00	%
Random Error :	0.09	μl	0.20	μl
Random Error :	0.18	%	0.40	%
Measurement Uncertainty :	\pm		0.14 μl	

Remarks

- The reported Expanded Uncertainty is calculated at 95.45 % C.L with coverage factor $k=2$
- The Above Results are within the maximum permissible Error

Calibrated by :


Ms. S. Manikala
(Calibration Engineer)



...redefining the true value



CERTIFICATE OF CALIBRATION

FT-Q-25		Page 1 of 2	
ULR No.	: CC29372000000801P	Date of Issue	: 21-Nov-20
Certificate No	: TVCS 20/11/189 - 01	Recom. Due Date	: 17-May-21
Date of Calibration	: 18-Nov-20		
Customer Details:		SRF No.	: 189
M/s. KALPANA MICROBIOLOGY LAB,		Calibrated at	: Lab
# 40/25, Pookadai Corner,		Date of Receipt	: 17-Nov-20
South Car St,		Cond. On Receipt	: Satisfactory
Thirucengode - 637211.			
Details of Test Instrument:			
Description	: Micropipette	Model No.	: --
Range	: 5-50 µl	Serial No	: YE174AB0061013
Least Count	: 0.5 µl	Identification No.	: --
Make	: D - Lab	Accuracy	: As per Manual
Working range	: --	Location	: Lab
Details of Standard Used			
Name	Certificate No.	Valid upto	Traceability
Electronic SemiMicro Balance	TVCSPL 20/05/364-01	12-May-21	TVCSPL, Chennai.
Work Instruction : WI-M-03 (Reference Standard ISO 8655-6)			
Environmental Details : Temperature : 25±2°C Relative Humidity : 50±10 % RH			
MECHANICAL CALIBRATION			
(Volume)			
Calibration Results			
1. Lower Volume :	5 µl	No. of Measurements :	10
<input type="text" value="4.99"/>	<input type="text" value="4.93"/>	<input type="text" value="4.98"/>	<input type="text" value="4.97"/>
<input type="text" value="5.00"/>	<input type="text" value="4.94"/>	<input type="text" value="4.96"/>	<input type="text" value="4.98"/>
<input type="text" value="5.00"/>	<input type="text" value="4.97"/>		
Mean Value :	4.98 µl		
Error Limits(±)			
Systematic Error :	-0.02 µl	0.13 µl	
Systematic Error :	-0.46 %	2.50 %	
Random Error :	0.02 µl	0.08 µl	
Random Error :	0.47 %	1.50 %	
Measurement Uncertainty :	±	0.14 µl	

Calibrated by :

S. Mani
Ms. S. Manikala
(Calibration Engineer)

Authorised by:

M. G. John
(The Head)



...redefining the true value



CERTIFICATE OF CALIBRATION

FT-Q-25		Page 1 of 2	
ULR No.	: CC29372000000802F	Date of Issue	: 21-Nov-20
Certificate No	: TVCS 20/11/189 - 02	Recom. Due Date	: 17-May-21
Date of Calibration	: 18-Nov-20	SRF No.	: 189
Customer Details:		Calibrated at	: Lab
M/s. KALPANA MICROBIOLOGY LAB,		Date of Receipt	: 17-Nov-20
# 40/25, Pookadai Corner,		Cond. On Receipt	: Satisfactory
South Car St,			
Thirucengode - 637211.			
Details of Test Instrument:			
Description	: Micropipette	Model	: --
Range	: 100 - 1000 µl	Serial No	: YE199AL0547646
Least Count	: 5 µl	Identification No.	: --
Make	: D - Lab	Accuracy	: As per Manual
Working range	: --	Location	: Lab
Details of Standard Used			
Name	Certificate No.	Valid upto	Traceability
Electronic SemiMicro Balance	TVCSPL 20/05/364 - 01	12-May-21	TVCSPL, Chennai.
Work Instruction	: WI-M-03 (Reference Standard ISO 8655-6)		
Environmental Details	: Temperature : 25±2°C Relative Humidity : 50±10 % RH		
MECHANICAL CALIBRATION			
(Volume)			
Calibration Results			
1. Lower Volume :	100 µl	No. of Measurements :	10
<input type="text" value="100.15"/>	<input type="text" value="100.22"/>	<input type="text" value="100.19"/>	<input type="text" value="100.17"/>
<input type="text" value="100.24"/>	<input type="text" value="100.19"/>	<input type="text" value="100.21"/>	<input type="text" value="100.25"/>
<input type="text" value="100.24"/>	<input type="text" value="100.22"/>		
Mean Value :	<input type="text" value="100.21"/> µl		
Error Limits(±)			
Systematic Error :	0.21 µl	0.80 µl	
Systematic Error :	0.21 %	0.80 %	
Random Error :	0.03 µl	0.30 µl	
Random Error :	0.03 %	0.30 %	
Measurement Uncertainty :	±	<input type="text" value="0.14"/> µl	

Calibrated by :
S. Mani
Ms. S. Manikala
(Calibration Engineer)

Authorised by:
(Signature)
Mr. G. John
(The Head)

...redefining the true value



ULR No. : CC29372000000802F
Certificate No : TVCS 20/11/189 - 02

MECHANICAL CALIBRATION (Volume) Calibration Results				
2. Middle Volume :	500	μl	No. of Measurements :	10
	499.63	499.59	499.56	499.60
	499.18	499.31	499.71	499.77
	499.56	499.26		
Mean Value :	499.51	μl		
Error Limits(\pm)				
Systematic Error :	-0.49	μl	4.00	μl
Systematic Error :	-0.10	%	0.80	%
Random Error :	0.20	μl	1.50	μl
Random Error :	0.04	%	0.30	%
Measurement Uncertainty :		\pm	1.1 μl	
3. Nominal Volume :	1000	μl	No. of Measurements :	10
	1000.51	999.50	999.73	999.64
	999.48	998.75	998.32	998.25
	998.47	999.06		
Mean Value :	999.17	μl		
Error Limits(\pm)				
Systematic Error :	-0.83	μl	8.00	μl
Systematic Error :	-0.08	%	0.80	%
Random Error :	0.73	μl	3.00	μl
Random Error :	0.07	%	0.30	%
Measurement Uncertainty :		\pm	1.1 μl	

Remarks

- The reported Expanded Uncertainty is calculated at 95.45 % C.L with coverage factor $k=2$
- The Above Results are within the maximum permissible Error

Calibrated by :

S. Mani
Ms. S. Manikala
(Calibration Engineer)



...redefining the true value