

## LAB MONTHLY SUMMARY

www.cmcwvilliers.ac.in/ajac/vw/monthlySummary.jsp?sample=1&cont=11&year=2020&group=7&yearGroup=68

Lab Name SANKALP LABORATORY  
 Month November  
 Constituent Group Chemistry II

Lab No 11338  
 Year 2020

Date of Result Entered : 04/11/2020

Date of Report Published : 05/12/2020

Sl.No Analyte Method / Principle Name

|   |                | Analyzer Name   | No of Participants | EV     | CV    | SD    | Your Value | SDS   | U    |
|---|----------------|---|--------------------|--------|-------|-------|------------|-------|------|
| 1 | GLUCOSE I      | GOD-POD   | 590                | 302.81 | 10.10 | 39.66 | 375        | -6.45 | 3.27 |
| 2 | CREATININE I   | JAFTE RATE BLANKED<br>JAFTE RATE (G) / ALBUMIN PIGRIB | 498                | 7.37   | 11.62 | 0.93  | 8 mg/dl    | 0.03  | 0.06 |
| 3 | T-PROTEIN I    | BIURET - colorimetric                                 | 424                | 5.34   | 9.16  | 0.49  | 5.3 g/dl   | -0.28 | 0.05 |
| 4 | ALBUMIN I      | BCG - colorimetric                                    | 462                | 3.23   | 9.31  | 0.30  | 3.5 g/dl   | 0.91  | 0.03 |
| 5 | CHOLESTEROL I  | CHOD-PAP  | 548                | 116.43 | 11.37 | 12.86 | 116        | -8.03 | 1.07 |
| 6 | TRIGLYCERIDE I | GPO-PAP / Enzymatic Colorimetric / End Point          | 525                | 148.27 | 15.04 | 22.30 | 199.88     | -4.15 | 1.95 |

| SDI Range            | Interpretation                           |
|----------------------|--|
| Within -1.0 to +1.0  | Excellent.                               |
| Between ±1.0 to ±2.0 | Good.                                    |
| Between ±2.0 to ±3.0 | Accept with caution, Warning Signal.     |
| Beyond ±3.0          | Unacceptable performance, Action Signal. |

Homogeneity and Stability of the sample is passed.

Data in CMC IQAS reports is confidential  
 Contact details:  
 Email: [clinfo@cmcvilliers.ac.in](mailto:clinfo@cmcvilliers.ac.in)  
 Contact Number: 0416-2283102

Page 1 of 1

*Pamela Christodou*  
 Dr. Pamela Christodou  
 CMC IQAS Co-Ordinator