

LAB MONTHLY SUMMARY

Lab Name **SHREE SANJEVE CLINICAL LABORATORY** Lab No **13700**
 Month **October** Year **2020**
 Constituent **Chemistry II**
 Group

Date of Result Entered : **19/10/2020**Date of Report Published : **09/11/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Biosystems	232	170.18	9.30	15.83	178 mg/dl	0.49	2.08
2	UREA I	UREASE BERTHELOT	Biosystems	60	113.64	9.95	11.31	110 mg/dl	-0.32	2.92
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Biosystems	209	4.71	11.56	0.55	4.8 mg/dl	0.17	0.08
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Biosystems	173	3.86	14.46	0.56	3.8 mg/dl	-0.11	0.08
5	T-PROTEIN I	BIURET - colorimetric	Biosystems	181	5.02	9.69	0.49	5.2 g/dl	0.37	0.07
6	ALBUMIN I	BCG - colorimetric	Biosystems	188	3.00	9.17	0.28	3 g/dl	0.00	0.04
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Biosystems	153	4.36	13.90	0.61	4.3 mg/dl	-0.10	0.10
8	CHOLESTEROL I	CHOD-PAP	Biosystems	206	98.79	10.78	10.65	101 mg/dl	0.21	1.48
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	189	189.02	10.67	20.16	195 mg/dl	0.30	2.93

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****