

20-11-2020

TO WHOMSOEVER IT MAY CONCERN

Dear Sir / Madam,

We are maintain everyday three times of our laboratory refrigerator temperature monitoring for Digital Thermometer monitor. We updating the record for every day. Reference record Refrigerator temperatures monitor chart, Format No F/MLL/10, Rev. No. 01, Date: 10-10-2020. We attached document. Kindly accept our request and make us favour.

Thank you

Yours Sincerely

Managing Partner

MEDLINE LABS
#75, SP Pudur, Paramathi Road,
NAMAKKAL - 637 001.
Cell : +91 94888 19000.



REFRIGERATOR TEMPERATURE MONITORING REPORT

FORMAT N:F/M/LL/10

REV NO: 01

REV.DATE:10-10-2020



S.No	Date	Time	Temperature °C	Signature	Year:	Time	Location:	Temperature °C	Signature	Refrigerator S.No:	Time	Temperature °C	Signature
1	01-11-2020	8.30 PM	6.3°C	[Signature]		12.00 PM	6.5°C	[Signature]			5. PM	6.1°C	[Signature]
2	02-11-2020	9.00 AM	T.1°C	[Signature]		12.00 PM	6.9°C	[Signature]			5.30 PM	T.0°C	[Signature]
3	03/11/2020	8.30 AM	6.2°C	[Signature]		12.30 PM	7.0°C	[Signature]			5. PM	5.1°C	[Signature]
4	04/11/2020	9.00 AM	5.3°C	[Signature]		1.00 PM	7.8°C	[Signature]			5. PM	6.0°C	[Signature]
5	05/11/2020	9.00 AM	2.9°C	[Signature]		12.30 PM	7.4°C	[Signature]			6.00 PM	5.1°C	[Signature]
6	06/11/2020	9.00 AM	T.1°C	[Signature]		12.30 PM	6.0°C	[Signature]			5.30 PM	3.9°C	[Signature]
7	07/11/2020	9.00 AM	T.8°C	[Signature]		12.00 PM	5.1°C	[Signature]			5.20 PM	2.8°C	[Signature]
8	08/11/2020	9.00 AM	T.9°C	[Signature]		12.00 PM	5.5°C	[Signature]			5.10 PM	4.6°C	[Signature]
9	09/11/2020	9.00 AM	6.5°C	[Signature]		12.00 PM	6.0°C	[Signature]			5.00 PM	5.6°C	[Signature]
10	10-11-20	8.30 AM	8.1°C	[Signature]		12.00 PM	6.1°C	[Signature]			6.30 PM	3.1°C	[Signature]
11	11-11-20	8.45 AM	T.0°C	[Signature]		12.30 PM	T.3°C	[Signature]			6.30 PM	2.8°C	[Signature]
12	12/11/2020	8.30 AM	2.8°C	[Signature]		12.00 PM	T.8°C	[Signature]			7.10 PM	3.3°C	[Signature]
13	13/11/2020	9.00 AM	6.1°C	[Signature]		1.00 PM	8.1°C	[Signature]			6.00 PM	3.6°C	[Signature]
14	14/11/2020	9.30 AM	6.5°C	[Signature]		1.00 PM	T.1°C	[Signature]			6.30 PM	4.0°C	[Signature]
15	15/11/2020	10.00 AM	6.9°C	[Signature]		1.00 PM	T.3°C	[Signature]			5.30 PM	5.1°C	[Signature]
16	16/11/2020	9.00 AM	7.1°C	[Signature]		12.00 PM	T.9°C	[Signature]			5. PM	5.4°C	[Signature]
17	17/11/2020	9.00 AM	8.0°C	[Signature]		12.00 PM	T.6°C	[Signature]			5. PM	6.1°C	[Signature]
18	18/11/2020	9.00 AM	7.0°C	[Signature]		12.00 PM	8.0°C	[Signature]			6.30 PM	T.3°C	[Signature]
19	19/11/2020	9.30 AM	6.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			5.30 PM	4.3°C	[Signature]
20	20/11/2020	10.18 AM	5.8°C	[Signature]		1.00 PM	7.6°C	[Signature]			5. PM	5.9°C	[Signature]
21	21/11/2020	8.00 AM	5.3°C	[Signature]		1.00 PM	T.6°C	[Signature]			6.00 PM	3.6°C	[Signature]
22	22/11/2020	8.80 AM	5.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			5.15 PM	2.5°C	[Signature]
23	23/11/2020	8.30 AM	5.1°C	[Signature]		1.00 PM	T.6°C	[Signature]			6.00 PM	4.8°C	[Signature]
24	24/11/2020	9.00 AM	5.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			6.00 PM	4.9°C	[Signature]
25	25/11/2020	8.00 AM	5.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			6.00 PM	4.8°C	[Signature]
26	26/11/2020	9.00 AM	5.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			6.00 PM	4.9°C	[Signature]
27	27/11/2020	9.00 AM	5.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			6.00 PM	4.8°C	[Signature]
28	28/11/2020	9.00 AM	5.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			6.00 PM	4.9°C	[Signature]
29	29/11/2020	9.00 AM	5.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			6.00 PM	4.8°C	[Signature]
30	30/11/2020	9.00 AM	5.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			6.00 PM	4.9°C	[Signature]
31	31/11/2020	9.00 AM	5.1°C	[Signature]		1.00 PM	T.0°C	[Signature]			6.00 PM	4.8°C	[Signature]

PREPARED BY NAME: [Signature] DATE: 10/10/2020

APPROVED BY NAME: [Signature] DATE: 10/10/2020

Mobile: 94888 19000

MEDLINE LABS

REFRIGERATOR TEMPERATURE MONITORING REPORT

FORMAT N: F/MILL/10

REV NO: 01

REV DATE: 10-10-2020



S.No	Date	Time	Month :		Signature	Time	Location :		Signature	Time	Temperature °C	Signature
			Year:	Year:			Temperature °C	Temperature °C				
1	1-12-20	9.00 am	3	12	W	12.00 pm	7.3°C	W	5.15 pm	6.4	W	
2	2-12-20	8.30 am	3	12	W	1 pm	7.9°C	W	5.30 pm	5.1	W	
3	3/12/20	8.00 am	2	12	W	1.30 pm	6.9°C	W	5.00 pm	4.0	W	
4	4-12-20	8.50 am	2	12	W	12.00 pm	8.1°C	W	6.00 pm	3.0	W	
5	5-12-20	8.30 am	2	12	W	12.00 pm	8.9°C	W	5.45 pm	5.3	W	
6	6/12/20	9.30 am	3	12	W	12.30 pm	8.1°C	W	7.00 pm	2.9	W	
7	7/12/20	9.00 am	2	12	W	1.00 pm	6.2°C	W	6.00 pm	4.1	W	
8	8/12/20	8.00 am	2	12	W	1.00 pm	7.1°C	W	7.00 pm	4.6	W	
9	9/12/20	8.00 am	2	12	W	1.30 pm	8.0°C	W	6.00 pm	6.1	W	
10	10/12/20	8.30 am	3	12	W	1.00 pm	7.6°C	W	5.30 pm	2.9	W	
11	11/12/20	9.00 am	2	12	W	1.30 pm	6.8°C	W	6.00 pm	5.2	W	
12	12/12/20	9.00 am	3	12	W	12.30 pm	7.3°C	W	5.40 pm	4.5	W	
13	13/12/20	8.00 am	3	12	W	12.45 pm	7.1°C	W	6.10 pm	3.9	W	
14	14/12/20	8.30 am	3	12	W	1.10 pm	8.9°C	W	6.20 pm	4.6	W	
15	15/12/20	9.00 am	3	12	W	1.15 pm	8.1°C	W	5.45 pm	6.1	W	
16	16/12/20	9.30 am	2	12	W	1.30 pm	8.2°C	W	5.35 pm	3.8	W	
17	17/12/20	8.30 am	2	12	W	1.25 pm	8.3°C	W	5.30 pm	3.9	W	
18	18/12/20	8.30 am	2	12	W	1.00 pm	7.8°C	W	5.45 pm	4.6	W	
19	19/12/20	8.30 am	2	12	W	1.00 pm	8.2°C	W	5.35 pm	6.1	W	
20	20/12/20	9.05 am	2	12	W	1.25 pm	8.3°C	W	5.30 pm	3.8	W	
21	21/12/20	9.05 am	2	12	W	1.00 pm	7.8°C	W	5.45 pm	4.6	W	
22												
23												
24												
25												
26												
27												
28												
29												
30												
31												

PREPARED BY NAME S Srinivasulu Reddy
SIGNATURE [Signature]
DATE 10-10-2020

APPROVED BY NAME [Signature]
SIGNATURE [Signature]
DATE 10-10-2020

Mobile: 9899 19000
 10-10-2020

Page 1 of 1

KANNAN A/C MASTER
(A/C & Refrigerator Service Center)

Siva Theatre Opposite,
ParamathiVelur,
Kabilarmalai Road,
Namakkal.
Pin-638182

P.Kannan BBA

Cell:94435 91398

Date : 18.11.2020

TO WHOMSOEVER IT MAY CONCERN

This is to certified that the Refrigerator in Medline Labs, Namakkal is Serviced and Calibration done, the working condition is good.

Next service due on 18.11.2021

Name of Model : LG
Model and Year : QL - B281BPZX/2018
Type : Direct Cool
Gress volume : 270 Liters
Storage Volume : 250 Liters
Colour : Gray

Instrument used for calibration : Temperature Calibrator 100mV to 1000mV

Date : 18.11.2020

For KANNAN A/c SALES & SERVICE

P. Kannan
Proprietor