

KANNAN A/C MASTER
(A/C & Refrigerator Service Center)

**Siva Theatre Opposite,
ParamathiVelur,
Kabilarmalai Road,
Namakkal.
Pin-638182**

P.Kannan BBA

Cell:94435 91398

SERVICE & CALIBRATION REPORT

This is to certified that the refrigerator in KALPANA MICROBIOLOGY LAB, Thiruchegode is service and calibration done, the I working condition is good

Customer Name : KALPANA MICROBIOLOGY LAB

Address: No.40/25, Pookadai Corner, South Car Street, Thiruchengode 637211

Name of model : Whirlpool

Model: A2070 399000A

Type: Direct Cool

Gross Volume : 310 Litters

Colour : Cherry Red

Instrument used for calibration : Temperature Calibrator

100mV to 1000mV

Calibration date :09-12-2020

Next calibration due date : 08-12-2021

For KANNAN A/c SALES & SERVICE

P. Kannan
Proprietor

REFRIGERATOR TEMPERATURE MONITORING REPORT

FORMAT N: F/KML/NABL/10

REV NO: 01

REV DATE: 10-12-2020

S.No	Date		Time		Temperature °C		Signature		Month: Nov 2020		Year: 2020		Location: Lab		Refrigerator S.No: 1		Temperature °C	Signature
	Date	Time	Time	Temperature °C	Signature	Time	Temperature °C	Signature	Time	Temperature °C	Time	Temperature °C	Signature	Time	Temperature °C	Signature		
1	01-11-2020	8:00 AM		6°C		8	12:00 PM		7.9°C		8	05:00 PM		6.2°C		8		
2	02-11-2020	8:20		A.E.		8	12:30 PM		6.5°C		8	5:00		7.0°C		8		
3	03-11-2020	8:30		A.8°C		8	12:40		6.5°C		8	5:15		7.2°C		8		
4	04-11-2020	8:00		A.5°C		8	12:20		A.8°C		8	5:40		5.4°C		8		
5	05-11-2020	8:10		5°C		8	12:40		5.5°C		8	5:30		6.6°C		8		
6	06-11-2020	9:00		A.5°C		8	12:15		A.8°C		8	5:25		6.8°C		8		
7	07-11-2020	7:50		A.7°C		8	12:20		6.0°C		8	5:30		7.4°C		8		
8	08-11-2020	8:00		5°C		8	12:45		6-8°C		8	5:40		6-3°C		8		
9	09-11-2020	8:15		A.6°C		8	12:20		6-6°C		8	5:15		5.9°C		8		
10	10-11-2020	8:20		A.1°C		8	12:15		7°C		8	5:20		6.0°C		8		
11	11-11-2020	9:00		A.8°C		8	12:00		5.9°C		8	5:30		7.0°C		8		
12	12-11-2020	9:00		5.2°C		8	12:35		6.6°C		8	5:00		6-8°C		8		
13	13-11-2020	8:45		3.9°C		8	12:40		6.8°C		8	5:15		6.6°C		8		
14	14-11-2020	8:00		A.3°C		8	12:10		7.0°C		8	6:20		6.4°C		8		
15	15-11-2020	8:30		5°C		8	12:30		7.3°C		8	5:40		7.0°C		8		
16	16-11-2020	8:35		A.8°C		8	12:10		6.8°C		8	5:55		7.2°C		8		
17	17-11-2020	8:50		5°C		8	12:50		6.4°C		8	5:00		6.6°C		8		
18	18-11-2020	9:00		5.0°C		8	12:30		6-6°C		8	5:10		6-2°C		8		
19	19-11-2020	8:10		A.3°C		8	12:10		6.8°C		8	6:40		6.4°C		8		
20	20-11-2020	8:25		A.8°C		8	12:20		7.0°C		8	5:30		7.2°C		8		
21	21-11-2020	8:30		A.1°C		8	12:25		7.2°C		8	5:40		6.4°C		8		
22	22-11-2020	8:40		A.5°C		8	12:35		6.6°C		8	5:10		7.2°C		8		
23	23-11-2020	8:20		A.8°C		8	12:10		6.9°C		8	5:30		6.8°C		8		
24	24-11-2020	8:25		3.7°C		8	12:15		7°C		8	5:20		6.9°C		8		
25	25-11-2020	8:00		2.8°C		8	12:50		7.2°C		8	5:40		7.1°C		8		
26	26-11-2020	9:00		A.3°C		8	12:10		6.8°C		8	5:10		6.6°C		8		
27	27-11-2020	8:15		A.8°C		8	12:15		6.6°C		8	5:00		5.9°C		8		
28	28-11-2020	8:10		A.1°C		8	12:55		7.2°C		8	5:30		5.8°C		8		
29	29-11-2020	8:30		A.5°C		8	12:10		6-7°C		8	5:40		6.6°C		8		
30	30-11-2020	8:40		A.8°C		8	12:20		7.3°C		8	5:10		6.8°C		8		

PREPARED BY NAME	SIPONITHA	APPROVED BY NAME	M. SADI ASQUAM
SIGNATURE	<i>[Signature]</i>	SIGNATURE	<i>[Signature]</i>
DATE	7-12-20	DATE	10-12-20

REFRIGERATOR TEMPERATURE MONITORING REPORT

REV NO: 01 REV DATE: 10-12-2020 Location: LAB

S.No	Date	Time	Temperature °C	Signature	Time	Temperature °C	Signature	Time	Temperature °C	Signature
1	1.12.20	8 AM	3.9°C	SP	12-PM	6.6°C	SP	5:00 PM	6.8°C	SP
2	2.12.20	8:10 AM	3.6°C	SP	12:30	6.7°C	SP	5:10	7.0°C	SP
3	3.12.20	8:15	3.7°C	SP	12:10	7.0°C	SP	5:15	7.2°C	SP
4	4.12.20	8:20	4.1°C	SP	12:40	6.3°C	SP	5:20	6.9°C	SP
5	5.12.20	8:30	4.8°C	SP	12:00	7.2°C	SP	5:00	6.4°C	SP
6	6.12.20	8:40	4.9°C	SP	12:15	6.4°C	SP	5:15	6.0°C	SP
7	7.12.20	8:10	5.0°C	SP	12:20	6.8°C	SP	5:30	7.3°C	SP
8	8.12.20	8:00	4.8°C	SP	12:10	7.0°C	SP	5:40	6.9°C	SP
9	9.12.20	8:15	5.4°C	SP	12:20	6.8°C	SP	5:10	7.4°C	SP
10	10.12.20	8:25	5.9°C	SP	12:30	6.9°C	SP	5:45	6.8°C	SP
11	11.12.20	8:40	4.8°C	SP	12:40	7.2°C	SP	5:35	6.6°C	SP
12	12.12.20	8:00	5.0°C	SP	12:10	5.9°C	SP	5:30	6.5°C	SP
13	13.12.20	8:15	5.4°C	SP	12:50	6.6°C	SP	5:20	7.2°C	SP
14	14.12.20	8:45	3.7°C	SP	12:30	6.8°C	SP	5:10	6.8°C	SP
15	15.12.20	8:30	4.8°C	SP	12:20	7.2°C	SP	5:20	7.5°C	SP
16	16.12.20	8:35	5.1°C	SP	12:15	6.8°C	SP	5:15	7.3°C	SP
17	17.12.20	8:20	4.9°C	SP	12:36	7.2°C	SP	5:25	6.8°C	SP
18	18.12.20	8:10	5.3°C	SP	12:40	7.5°C	SP	5:30	6.5°C	SP
19	19.12.20	8:00	5.4°C	SP	12:00	7.0°C	SP	5:00	5.9°C	SP
20	20.12.20	8:15	5.5°C	SP	12:30	7.3°C	SP	5:40	6.4°C	SP
21	21.12.20	7:50	4.9°C	SP	12:40	6.8°C	SP	5:00		
22										
23										
24										
25										
26										
27										
28										
29										
30										
31										

PREPARED BY NAME	S.PUNITHA	APPROVED BY NAME	M-SAD HASILUAM
SIGNATURE		SIGNATURE	
DATE		DATE	



REFRIGERATOR TEMPERATURE MONITORING REPORT

Kajman Microbiology Lab REV NO: 01 REV DATE: 10-12-2020

FORMAT N- F/KML/NABL/10 Month: 11/2020 Year: 2020 Location: Lab Refrigerator S.No: 1

S.No	Date	Time	Temperature °C	Signature	Time	Temperature °C	Signature	Time	Temperature °C	Signature
1	11-11-20	8:00 AM	6.2°C	[Signature]	12:00 PM	7.9°C	[Signature]	05:00 PM	6.2°C	[Signature]
2	11-11-20	8:00 AM	6.2°C	[Signature]	12:30 PM	6.5°C	[Signature]	5:00 PM	7.0°C	[Signature]
3	11-11-20	8:30 AM	6.5°C	[Signature]	1:00 PM	6.5°C	[Signature]	5:15 PM	7.2°C	[Signature]
4	11-11-20	9:00 AM	6.8°C	[Signature]	12:30 PM	6.8°C	[Signature]	5:30 PM	7.2°C	[Signature]
5	11-11-20	9:10 AM	5.9°C	[Signature]	12:40 PM	5.5°C	[Signature]	5:35 PM	6.8°C	[Signature]
6	11-11-20	9:20 AM	6.2°C	[Signature]	12:45 PM	6.2°C	[Signature]	5:40 PM	7.2°C	[Signature]
7	11-11-20	9:30 AM	6.2°C	[Signature]	12:50 PM	6.2°C	[Signature]	5:45 PM	6.3°C	[Signature]
8	11-11-20	9:40 AM	5.9°C	[Signature]	1:00 PM	6.2°C	[Signature]	5:50 PM	5.9°C	[Signature]
9	11-11-20	9:55 AM	6.2°C	[Signature]	1:05 PM	6.2°C	[Signature]	5:55 PM	6.0°C	[Signature]
10	11-11-20	10:10 AM	6.2°C	[Signature]	1:15 PM	7.0°C	[Signature]	6:00 PM	6.5°C	[Signature]
11	11-11-20	10:20 AM	6.2°C	[Signature]	1:20 PM	5.9°C	[Signature]	6:05 PM	6.0°C	[Signature]
12	11-11-20	10:30 AM	5.2°C	[Signature]	1:25 PM	6.2°C	[Signature]	6:10 PM	6.6°C	[Signature]
13	11-11-20	10:45 AM	5.9°C	[Signature]	1:30 PM	6.2°C	[Signature]	6:15 PM	6.6°C	[Signature]
14	11-11-20	10:55 AM	6.2°C	[Signature]	1:35 PM	7.0°C	[Signature]	6:20 PM	6.4°C	[Signature]
15	11-11-20	11:00 AM	6.2°C	[Signature]	1:40 PM	7.3°C	[Signature]	6:25 PM	7.2°C	[Signature]
16	11-11-20	11:05 AM	6.2°C	[Signature]	1:45 PM	6.8°C	[Signature]	6:30 PM	7.2°C	[Signature]
17	11-11-20	11:10 AM	5.9°C	[Signature]	1:50 PM	6.8°C	[Signature]	6:35 PM	6.6°C	[Signature]
18	11-11-20	11:15 AM	6.2°C	[Signature]	1:55 PM	6.8°C	[Signature]	6:40 PM	6.6°C	[Signature]
19	11-11-20	11:20 AM	6.2°C	[Signature]	2:00 PM	6.8°C	[Signature]	6:45 PM	6.6°C	[Signature]
20	11-11-20	11:25 AM	6.2°C	[Signature]	2:05 PM	7.2°C	[Signature]	6:50 PM	7.2°C	[Signature]
21	11-11-20	11:30 AM	6.2°C	[Signature]	2:10 PM	7.2°C	[Signature]	6:55 PM	7.2°C	[Signature]
22	11-11-20	11:35 AM	6.2°C	[Signature]	2:15 PM	7.2°C	[Signature]	7:00 PM	6.8°C	[Signature]
23	11-11-20	11:40 AM	6.2°C	[Signature]	2:20 PM	6.6°C	[Signature]	7:05 PM	6.8°C	[Signature]
24	11-11-20	11:45 AM	6.2°C	[Signature]	2:25 PM	6.6°C	[Signature]	7:10 PM	6.8°C	[Signature]
25	11-11-20	11:50 AM	6.2°C	[Signature]	2:30 PM	7.2°C	[Signature]	7:15 PM	6.8°C	[Signature]
26	11-11-20	11:55 AM	6.2°C	[Signature]	2:35 PM	7.2°C	[Signature]	7:20 PM	6.8°C	[Signature]
27	11-11-20	12:00 PM	6.2°C	[Signature]	2:40 PM	6.6°C	[Signature]	7:25 PM	6.8°C	[Signature]
28	11-11-20	12:05 PM	6.2°C	[Signature]	2:45 PM	7.2°C	[Signature]	7:30 PM	6.8°C	[Signature]
29	11-11-20	12:10 PM	6.2°C	[Signature]	2:50 PM	6.7°C	[Signature]	7:35 PM	6.8°C	[Signature]
30	11-11-20	12:15 PM	6.2°C	[Signature]	2:55 PM	7.2°C	[Signature]	7:40 PM	6.8°C	[Signature]
31	11-11-20	12:20 PM	6.2°C	[Signature]	3:00 PM	7.2°C	[Signature]	7:45 PM	6.8°C	[Signature]

APPROVED BY NAME: A. SATHYANARAYAN DATE: 10-12-2020

PREPARED BY NAME: S. P. J. THA. SIGNATURE: [Signature] DATE: 7-12-2020

Page 1 of 1