



PC-1034

CMC VIROEQAS
By National Reference Laboratory (NRL)
Under National AIDS Control Organisation (NACO)
PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmevellore.ac.in Phone: 0416-2283455

PANEL: HIV Serology

Name of the laboratory: ICTC THQH KOYILANDY

Distribution Round: Round -1, 2020-21

Panel dispatch date: 07/10/2020

Result received month: October 2020

SL NO	SAMPLE ID	INTENDED RESULT	YOUR RESULT	MATCHING YES / NO	% MATCHING
1	S012001	POSITIVE FOR HIV-1	POSITIVE FOR HIV-1	YES	100
2	S012002	NEGATIVE	NEGATIVE	YES	100
3	S012003	NEGATIVE	NEGATIVE	YES	100
4	S012004	POSITIVE FOR HIV-1	POSITIVE FOR HIV-1	YES	100

REMARKS: PERFORMANCE IS SATISFACTORY

Comments:

Participants are scored based on qualitative result.

PT Unit, Department of Clinical Virology, Christian Medical College, Vellore

The data in this CMCVIROEQAS reports are confidential



PC-1034

CMC VIROEQAS
By National Reference Laboratory (NRL)
Under National AIDS Control Organisation (NACO)
PT Unit, Department of Clinical Virology, Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 0416-2283455

Assigned Value:

Assigned value is determined using multiple assays/multiple testing of the same analyte. The proficiency testing material is a pooled lyophilized plasma sample prepared by following institutional proficiency testing standard operating protocol (PT-SOP) by trained staff. None of the work related to testing, preparation and packaging of sample is subcontracted. The homogeneity of the PT material is determined using multiple testing of multiple aliquots. The stability of the testing material is determined by assessing the reactivity of the specimen till the closing date.

Enquires: For queries, please contact CMCVIROEQAS coordinator at email: nrl.cmc@gmail.com

Name of CMCVIROEQAS coordinator

Signature

Dr. Rajesh Kannangai

Department of Clinical Virology

Report Dispatch Date: 22/12/2020

Report authorized by: CMCVIROEQAS Coordinator

END OF REPORT

PT Unit, Department of Clinical Virology, Christian Medical College, Vellore
The data in this CMCVIROEQAS reports are confidential