

## CALIBRATION REPORT

**Details of the Unit Under Calibration:**

**Calibration Certificate No: SBS/CL/20-21/00907**

Page: 1 of 2

<b>Customer Name:</b>	JOTHI X-RAY ECG & LABORATORY	<b>Customer Address:</b>	1, SIVAN KOVIL STREET, THOOTHUKUDI-628002.
<b>Location:</b>	BIOCHEMISTRY	<b>Serial No:</b>	315159
<b>Description:</b>	MICRO PIPETTE	<b>Tag/Asset No:</b>	NA
<b>Make:</b>	MERILETTE	<b>Date of Calibration:</b>	20-09-2020
<b>Model:</b>	MERIL	<b>Date of Due:</b>	19-09-2021
<b>Range:</b>	100-1000µl		VARIABLE

**Traceability Details:**

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021


**Environmental Conditions and Standard Operating Procedure Details:**

ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS			
Temperature	24°C	Calibration Done at	ONSITE
Humidity(40-70% RH)	55 %	Calibration Procedure No	--

**REMARKS:**

Test Report : Attached

Calibrated by:



BALAJI.M

(Calibration Engineer)

Approved by:



C.SHANMUGARAJ  
(Technical Manager)

Result of Calibration: SBS/TR/20-21/00907

Page: 2 of 2

TEST REPORT				
Units	Nominal Values	Measured Values	Tolerance	Result
$\mu$ l	100	100.2	$\pm 0.5$ %	PASS
$\mu$ l	300	300.1	$\pm 0.5$ %	PASS
$\mu$ l	1000	1000.0	$\pm 0.5$ %	PASS



**CALIBRATION RESULT:**

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: 

Date: 20-09-2020