

CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/00909

Page: 1 of 2

Customer Name:	JOTHI X-RAY ECG & LABORATORY	Customer Address:	1, SIVAN KOVIL STREET, THOOTHUKUDI-628002.
Location:	BIOCHEMISTRY	Serial No:	YE18AAL0147045
Description:	MICRO PIPETTE	Tag/Asset No:	NA
Make:	DRAGON	Date of Calibration:	20-09-2020
Model:	NA	Date of Due:	19-09-2021
Range:	100-1000µl	VARIABLE	

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021

Environmental Conditions and Standard Operating Procedure Details:

<i>ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS</i>			
Temperature	24°C	Calibration Done at	ONSITE
Humidity(40-70% RH)	55 %	Calibration Procedure No	--

REMARKS:

Test Report : Attached

Calibrated by:



BALAJIM

(Calibration Engineer)

Approved by:



C.SHANMUGARAJ
(Technical Manager)

Result of Calibration: SBS/TR/20-21/00909

Page: 2 of 2

TEST REPORT				
Units	Nominal Values	Measured Values	Tolerance	Result
μ l	100	100.0	$\pm 0.5 \%$	PASS
μ l	500	500.1	$\pm 0.5 \%$	PASS
μ l	1000	1000.2	$\pm 0.5 \%$	PASS



CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: 

Date: 20-09-2020