

LAB MONTHLY SUMMARY

Lab Name SRI ANNAI CLINICAL LAB

Lab No 4869

Month October

Year 2020

Constituent Chemistry II
GroupDate of Result Entered :
13/10/2020Date of Report Published :
09/11/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Transasia / Erba	654	169.94	10.68	18.14	191.4 mg/dl	1.18	1.42
2	UREA I	UREASE UV / GLDH	Transasia / Erba	552	112.30	13.65	15.32	132.1 mg/dl	1.29	1.30
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Transasia / Erba	507	4.67	11.38	0.53	6.3 mg/dl	3.07	0.05
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) / JENDRASSIK	Transasia / Erba	503	3.61	12.89	0.47	3.9 mg/dl	0.62	0.04
5	T-PROTEIN I	BIURET - colorimetric	Transasia / Erba	459	5.18	8.64	0.45	5.9 g/dl	1.61	0.04
6	ALBUMIN I	BCG - colorimetric	Transasia / Erba	459	3.05	8.97	0.27	3.4 g/dl	1.28	0.03
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Transasia / Erba	467	4.33	17.71	0.77	6 mg/dl	2.17	0.07
8	CHOLESTEROL I	CHOD-PAP	Transasia / Erba	555	96.92	13.51	13.09	105.3 mg/dl	0.64	1.11
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	530	168.92	12.76	21.55	182.3 mg/dl	0.62	1.87
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ± 1.0 to ± 2.0		Good.								
Between ± 2.0 to ± 3.0		Accept with caution. Warning Signal.								
Beyond ± 3.0		Unacceptable performance. Action Signal.								

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*Dr. Pamela Christudoss
CMC EQAS Co-Ordinator



Sri Annai Clinical Lab

X-Ray & E.C.G.

☎ : 224848

CORRECTIVE ACTION OF OCTOBER 2020 EQAS RESULT

Date 11.11.2020

Creatinine and Uric acid values are high in October 2020


CORRECTIVE ACTION :

New Kit Changed and Proper washing on Machine and clean the flow cell.

Proper standard run on Creatinine and Uric acid

Internal QC Done. Internal QC Values on Creatinine and Uric acid with in QC Ranges.

Problems Solved.


SRI ANNAI CLINICAL LAB & X-RAY
11, PATTABI RAMAR KOVIL STREET
ARUPPUKOTTAI - 626 101