



International Laboratory for
Metrology
BIPM
10 Avenue des Bories
92310 Suresnes Cedex
France
Tel: +33 (0)1 47 90 60 01
Fax: +33 (0)1 47 90 60 02
E-mail: metro@bipm.fr
www.bipm.fr

Erba
METTECH

CHEM-5 Plus v2



Ref No. TBMSZ1/CAL/0558

Date : 22th Dec' 2020

IDHAYAM DIAGNOSTIC SERVICE
No.26, Dr.R.P.ROAD,
K.K.NAGAR,
HASTHINAPURAM,
CHROMPET,
CHENNAI-600 044.

Dear Sirs,

As desired by you please find the Calibration Check Report on EC-5 plusV2 Sr No. N120575

1. Year of manufacturing: 2012
2. Range of the instrument: Optical density Linear form 0-2.5
3. Date of Calibration: 22th Dec '2020
4. Next Calibration Due: 21th Dec '2021

Test Voltages / Test Temperature -

Test Voltage	Actual Voltage	Ok / Not Ok
Logic 5 V	5.01V	OK
Line voltage	227V	OK
Lamp Voltage 12v	11.67V	OK
Temperature 25°C	25°C	OK
Temperature 30°C	30°C	OK
Temperature 37°C	37°C	OK
Input Voltage	18.0 VDC	OK
Pump Calibration	1766	OK

Transasia Bio - Medicals Ltd., No: 8/3, Unique Sterling, 3rd Cross Street, Sterling Road, Nungambakkam,
Chennai - 600034 Tel: +91 44 3297 7703/ 2822 7149 Fax: +91 44 2825 9435 Email: sz1@transasia.co.in
CIN: U33110MH1985PLC036198

TRANSASIA®



UNMATCHED SERVICE
SINCE 1979...

Sr.No.	Filters	Pre Channel	Log Channel
1	340	1.44L	0.07
2	405	1.47L	0.06
3	450	1.48L	0.06
4	505	1.44L	0.06
5	546	1.42L	0.07
6	578	1.40L	0.07
7	630	1.41L	0.07
8	670	1.37L	0.09

Thanking you and assuring of our prompt attention at all times.

Yours Sincerely,
For TRANSASIA BIO-MEDICALS LTD.,

S.SIVA
SERVICE ENGINEER