

LAB MONTHLY SUMMARY

Lab Name SRI ANNAI CLINICAL LAB

Lab No 4869

Month November

Year 2020

Constituent Chemistry II
GroupDate of Result Entered :
04/11/2020Date of Report Published :
05/12/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Transasia / Erba	590	392.81	10.10	39.66	425 mg/dl	0.81	3.27
2	UREA I	UREASE UV / GLDH	Transasia / Erba	515	58.97	10.38	6.12	72.4 mg/dl	2.19	0.54
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Transasia / Erba	498	7.97	11.62	0.93	9.4 mg/dl	1.54	0.08
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Transasia / Erba	495	1.69	18.36	0.31	1.65 mg/dl	-0.13	0.03
5	T-PROTEIN I	BIURET - colorimetric	Transasia / Erba	424	5.34	9.16	0.49	6 g/dl	1.35	0.05
6	ALBUMIN I	BCG - colorimetric	Transasia / Erba	462	3.23	9.21	0.30	3.3 g/dl	0.23	0.03
7	CHOLESTEROL I	CHOD-PAP	Transasia / Erba	548	110.43	11.37	12.56	190.5 mg/dl	6.37	1.07
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	525	148.27	15.04	22.30	149.9 mg/dl	0.07	1.95

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College,
Vellore

***** End of Report *****



Sri Annai Clinical Lab

X-Ray & E.C.G.

☎ : 224848

CORRECTIVE ACTION OF NOVEMBER 2020 EQAS RESULT

Date 10-12-2020

Urea and Cholesterol values are high in November 2020

CORRECTIVE ACTION :

New Kit Changed and Proper washing on Machine and clean the flow cell.

Proper standard run on Urea and Cholesterol

Internal QC Done. Internal QC Values on Urea and Cholesterol with in QC Ranges.

Problems Solved.

SRI ANNAI CLINICAL LAB & X-RAY
11, PATTABI RAMAR KOVIL STREET
ARUPPUKOTTAI - 626 101