



Customer Details :	Qualitycare Diagnostic Centre 165, Kallar Society Building Madurai Road, THENI.
Instrument Details:	ACCUREX SPHERA Sr. No. 172207002

**CALIBRATION CERTIFICATE FOR FULLY AUTO MATATED ANALYZER**

Dear sir/Madam,

We have carried out the calibration of SPHERA Biochemistry FULLY AUTOMATED ANALYZER (Sr.No- 172207002) Belonging to your biochemistry Laboratory on 02-11-2020 the calibration values obtained are as follows.

**PHOTOMETER CHECK**

PREVIOUS DATA			CURRENT DATA		
WAVELENGTH	mVolt	Efficiency(%)	WAVELENGTH	mVolt	Efficiency(%)
340	1480.4	100	340	1502.5	100
405	1448.2	97	405	1548.6	100
505	1469.4	96	505	1452.2	100
546	1455.2	99	546	1415.8	101
578	1477.7	100	578	1444.6	101
600	1481.3	99	600	1489.7	100
650	1466.4	100	650	1548.4	100
700	1498.9	100	700	1514.1	100
0	0	0	0	0	0

The above mentioned values are found to be within the standard specified by our principles, ACCUREX BIOMEDICAL PVT LTD, MUMBAI.

The next calibration due for this analyzer is due on 01-11-2021

Thanking you,  
Yours truly,

For ACCUREX BIOMEDICAL PVT LTD.

  
SENTHIL KUMAR.R  
South Service Manager.

