



### CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THQH/REF/D01	Date of Issue : 23/11/2020
		Calibration Done On : 21/11/2020
		Calibration Due On : 20/11/2021

Customer Name	TALUK HEAD QUARTERS HOSPITAL, HARIPAD.
---------------	--

#### DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	Haier	-	80042447	ICTC Lab

#### DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/102669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

<b>Environmental Conditions</b>	Temperature	27 °C	Relative Humidity	55 %
---------------------------------	-------------	-------	-------------------	------

#### PERFORMANCE DATA

Display	Measured	Tolerance
1.3 °C	1.2 °C	± 0.3 °C
2.4 °C	2.4 °C	± 0.3 °C
4.2 °C	4.3 °C	± 0.3 °C
7.2 °C	7.2 °C	± 0.3 °C

<b>Calibration Status</b>	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

Approved By

Prabin KP (Service Engineer)

Binoj Thomas (Quality Manager)