



## CERTIFICATE OF CALIBRATION

Certification No: MES/201126/C1

Calibrated for: THALUK HEAD QUARTERS HOSPITAL, ICTC LAB, HARIPAD.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 10-100 $\mu$ l

SL.NO/ID : NA/80042448

NO OF CHANNELS : 1

Date of Calibration : 26/11/2020

Calibrated at : Permanent Facility

MAKE : DR.PETTE

MODEL NO : NA

Resolution of DUC: 1  $\mu$ l

Next calibration due : 25/11/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 26/11/2020

Calibrated by :

PRABIN. K P

(Lab In Charge )

Approved Signatory

BINOJ THOMAS  
Trin  
(Quality / Technical Manager)

## CALIBRATION REPORT

Certification No: MES/201126/C1

### CALIBRATION RESULT OF MICROPIPETTE

Sl. No.	Set Volume $\mu\text{l}$	Actual Volume $\mu\text{l}$	# Expanded uncertainty ( $\mu\text{l}$ )	Coverage factor (k)
1	10	9.95	$\pm 0.41$	2
2	50	49.96	$\pm 0.409$	2
3	100	99.99	$\pm 0.409$	2

# The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

#### Environmental Factors:

Temperature:  $26.5^{\circ}\text{C} \pm 1^{\circ}\text{C}$

Air pressure: 1003 mbar

Double distilled water,

Relative humidity: 55 %

#### **Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration **due date** mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Calibration is carried out as per MES work procedure: MES/WP/VOLUME.