

LAB MONTHLY SUMMARY

Lab Name **Marson Clinical Laboratory** Lab No **11636**
 Month **August** Year **2020**
 Constituent Group **Chemistry II**

Date of Result Entered : 06/08/2020

Date of Report Published : 10/09/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	819	111.85	10.06	11.25	120 mg/dl	0.72	0.79
2	UREA I	Urease Berthelot	Any Analyser	247	71.64	14.78	10.59	83 mg/dl	1.07	1.35
3	CREATININE I	Enzymatic Colorimetric	Any Analyser	78	9.06	13.70	1.24	8.31 mg/dl	-0.60	0.28
4	T.BILIRUBIN I	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser	531	5.89	14.10	0.83	7.2 mg/dl	1.58	0.07
5	T-PROTEIN I	Biuret - Colorimetric	Any Analyser	630	5.35	10.21	0.55	4.91 g/dl	-0.81	0.04
6	ALBUMIN I	BCG - colorimetric	Any Analyser	579	3.20	9.46	0.30	3.27 g/dl	0.23	0.03
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Any Analyser	535	5.96	15.02	0.90	5.37 mg/dl	-0.66	0.08
8	CHOLESTEROL I	CHOD-PAP	Any Analyser	647	109.71	11.44	12.55	117 mg/dl	0.58	0.99
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	622	270.58	11.97	32.39	274 mg/dl	0.11	2.60

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****