

LAB MONTHLY SUMMARY

Lab Name **Marson Clinical Laboratory** Lab No **11636**
 Month **September** Year **2020**
 Constituent Group **Chemistry II**

Date of Result Entered : 05/09/2020

Date of Report Published : 07/10/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	838	223.99	10.76	24.09	232.7 mg/dl	0.36	1.66
2	UREA I	Urease Berthelot	Any Analyser	255	122.58	15.76	19.32	143 mg/dl	1.06	2.42
3	CREATININE I	Enzymatic Colorimetric	Any Analyser	82	1.42	13.55	0.19	1.31 mg/dl	-0.57	0.04
4	T.BILIRUBIN I	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser	588	5.24	14.11	0.74	6.44 mg/dl	1.62	0.06
5	T-PROTEIN I	Biuret - Colorimetric	Any Analyser	666	5.41	11.03	0.60	4.83 g/dl	-0.97	0.05
6	ALBUMIN I	BCG - colorimetric	Any Analyser	603	3.23	9.57	0.31	3.27 g/dl	0.13	0.03
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Any Analyser	645	7.12	15.23	1.08	6.53 mg/dl	-0.54	0.09
8	CHOLESTEROL I	CHOD-PAP	Any Analyser	726	108.64	11.67	12.68	114.6 mg/dl	0.47	0.94
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	890	99.84	14.79	14.77	90.3 mg/dl	-0.65	0.99

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****