

LAB MONTHLY SUMMARY

Lab Name **Marson Clinical Laboratory** Lab No **11636**
 Month **October** Year **2020**
 Constituent Group **Chemistry II**

Date of Result Entered : 13/10/2020

Date of Report Published : 09/11/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	910	168.35	10.59	17.82	190.4 mg/dl	1.24	1.18
2	UREA I	Urease Berthelot	Any Analyser	284	104.04	17.45	18.16	120 mg/dl	0.88	2.15
3	CREATININE I	Enzymatic Colorimetric	Any Analyser	98	4.78	12.80	0.61	4.43 mg/dl	-0.57	0.12
4	T.BILIRUBIN I	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser	614	3.79	15.81	0.60	4.42 mg/dl	1.05	0.05
5	T-PROTEIN I	Biuret - Colorimetric	Any Analyser	667	5.06	10.46	0.53	4.33 g/dl	-1.38	0.04
6	ALBUMIN I	BCG - colorimetric	Any Analyser	622	3.02	9.51	0.29	2.96 g/dl	-0.21	0.02
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Any Analyser	517	4.49	18.56	0.83	4.57 mg/dl	0.10	0.07
8	CHOLESTEROL I	CHOD-PAP	Any Analyser	771	101.13	13.52	13.68	111.1 mg/dl	0.73	0.98
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	735	180.95	13.56	24.53	185.4 mg/dl	0.18	1.81

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****