

## LAB MONTHLY SUMMARY

Lab Name **Marson Clinical Laboratory** Lab No **11636**  
 Month **November** Year **2020**  
 Constituent Group **Chemistry II**

Date of Result Entered : 10/11/2020

Date of Report Published : 05/12/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	912	373.63	10.97	40.98	411.76 mg/dl	0.93	2.71
2	UREA I	Urease UV / GLDH	Any Analyser	477	57.91	10.73	6.21	63.29 mg/dl	0.87	0.57
3	CREATININE I	Enzymatic Colorimetric	Any Analyser	92	8.14	10.57	0.86	8 mg/dl	-0.16	0.18
4	T.BILIRUBIN I	OTHERS ( Any Other Principles / Methods )	Any Analyser	163	1.77	19.89	0.35	2.96 mg/dl	3.4	0.06
5	T-PROTEIN I	Biuret - Colorimetric	Any Analyser	661	5.24	9.34	0.49	4.83 g/dl	-0.84	0.04
6	ALBUMIN I	BCG - colorimetric	Any Analyser	641	3.19	9.75	0.31	3.11 g/dl	-0.26	0.02
7	CHOLESTEROL I	CHOD-PAP	Any Analyser	767	113.22	11.86	13.43	110.25 mg/dl	-0.22	0.97
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	723	159.55	13.88	22.15	151.34 mg/dl	-0.37	1.65

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*