

LAB MONTHLY SUMMARY

Lab Name **Marson Clinical Laboratory** Lab No **11636**
 Month **December** Year **2020**
 Constituent Group **Chemistry II**

Date of Result Entered : 12/12/2020

Date of Report Published : 31/12/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	879	143.64	10.05	14.43	155 mg/dl	0.79	0.97
2	UREA I	Urease UV / GLDH	Any Analyser	503	80.76	13.55	10.94	94.8 mg/dl	1.28	0.98
3	CREATININE I	Enzymatic Colorimetric	Any Analyser	105	4.02	11.89	0.48	4.15 mg/dl	0.27	0.09
4	T.BILIRUBIN I	OTHERS (Any Other Principles / Methods)	Any Analyser	163	2.18	21.31	0.46	2.36 mg/dl	0.39	0.07
5	T-PROTEIN I	Biuret - Colorimetric	Any Analyser	672	4.91	10.37	0.51	5.65 g/dl	1.45	0.04
6	ALBUMIN I	BCG - colorimetric	Any Analyser	629	3.01	10.24	0.31	3.19 g/dl	0.58	0.02
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Any Analyser	565	4.29	19.63	0.84	4.9 mg/dl	0.72	0.07
8	CHOLESTEROL I	CHOD-PAP	Any Analyser	704	101.54	11.40	11.57	106.07 mg/dl	0.39	0.87
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	708	107.44	16.94	18.20	97.4 mg/dl	-0.55	1.37

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****