

## LAB MONTHLY SUMMARY

Lab Name **Marson Clinical Laboratory** Lab No **11636**  
 Month **July** Year **2020**  
 Constituent Group **Chemistry II**

Date of Result Entered : 16/07/2020

Date of Report Published : 14/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	817	379.38	11.69	44.35	421 mg/dl	0.94	3.10
2	UREA I	Urease Berthelot	Any Analyser	239	87.31	17.09	14.92	101 mg/dl	0.92	1.93
3	CREATININE I	Enzymatic Colorimetric	Any Analyser	91	0.96	22.61	0.22	0.81 mg/dl	-0.69	0.05
4	T.BILIRUBIN I	Diazonium salt ( Colorimetric ) / Jendrassik	Any Analyser	553	4.23	16.70	0.71	5.18 mg/dl	1.35	0.06
5	T-PROTEIN I	Biuret - Colorimetric	Any Analyser	623	5.37	10.23	0.55	4.93 g/dl	-0.80	0.04
6	ALBUMIN I	BCG - colorimetric	Any Analyser	599	3.25	9.69	0.32	3.57 g/dl	1.02	0.03
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Any Analyser	582	7.52	13.86	1.04	6.04 mg/dl	-1.42	0.09
8	CHOLESTEROL I	CHOD-PAP	Any Analyser	728	108.74	12.51	13.60	121 mg/dl	0.90	1.01
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	705	212.56	13.62	28.94	200 mg/dl	-0.43	2.18

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1


Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*