

## LAB MONTHLY SUMMARY

Lab Name **Marson Clinical Laboratory**Lab No **11636**Month **June**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **17/06/2020**Date of Report Published : **08/08/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	845	80.95	11.31	9.16	87.3 mg/dl	0.69	0.63
2	UREA I	Urease Berthelot	Any Analyser	257	48.27	16.28	7.86	60 mg/dl	1.49	0.98
3	CREATININE I	Enzymatic Colorimetric	Any Analyser	101	3.54	11.48	0.41	3.2 mg/dl	-0.84	0.08
4	T.BILIRUBIN I	Diazonium salt ( Colorimetric ) / Jendrassik	Any Analyser	635	5.39	17.93	0.97	6.42 mg/dl	1.07	0.08
5	T-PROTEIN I	Biuret - Colorimetric	Any Analyser	663	4.87	11.36	0.55	4.82 g/dl	-0.09	0.04
6	ALBUMIN I	BCG - colorimetric	Any Analyser	633	2.98	10.47	0.31	3.2 g/dl	0.71	0.02
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Any Analyser	561	4.60	21.29	0.98	3.88 mg/dl	-0.74	0.08
8	CHOLESTEROL I	CHOD-PAP	Any Analyser	790	97.69	15.59	15.23	108.5 mg/dl	0.71	1.08
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	710	87.69	20.45	17.94	88.6 mg/dl	0.05	1.35

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102



Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*