

## LAB MONTHLY SUMMARY

Lab Name **Marson Clinical Laboratory** Lab No **11636**  
 Month **May** Year **2020**  
 Constituent Group **Chemistry II**

Date of Result Entered : 17/06/2020

Date of Report Published : 04/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	858	105.04	10.46	10.98	123.5 mg/dl	1.68	0.75
2	UREA I	Urease Berthelot	Any Analyser	226	29.17	17.27	5.04	35 mg/dl	1.16	0.67
3	CREATININE I	Enzymatic Colorimetric	Any Analyser	80	4.71	12.68	0.60	4.3 mg/dl	-0.69	0.13
4	T.BILIRUBIN I	Diazonium salt ( Colorimetric ) / Jendrassik	Any Analyser	581	2.89	18.87	0.55	3.51 mg/dl	1.14	0.05
5	T-PROTEIN I	Biuret - Colorimetric	Any Analyser	639	5.23	9.21	0.48	4.84 g/dl	-0.81	0.04
6	ALBUMIN I	BCG - colorimetric	Any Analyser	635	3.20	10.09	0.32	3.49 g/dl	0.90	0.03
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Any Analyser	609	6.34	14.77	0.94	5 mg/dl	-1.43	0.08
8	CHOLESTEROL I	CHOD-PAP	Any Analyser	724	108.65	11.96	12.99	120 mg/dl	0.87	0.97
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	652	235.57	12.43	29.29	233 mg/dl	-0.09	2.29

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*