

CALIBRATION CERTIFICATE

This is to certify that the following Biochemistry Analyser has been tested on the below aspects.

Customer Name : ALAKNANDA DIAGNOSTICS LAB
Location : H 1, Vikrmadjtya Tower
 Alaknanda Shopping Complex
 Alaknanda
 New Delhi, NORTH 110019
Instrument : TURBOCHEM 100 (4611) CHEMISTRY AUTO ANALYZER
Installation Date : 21-01-2013
SerialNo : 4611-1086
Calibrated On : 21-12-2020
Next Calibration Due On : 21-12-2021

FILTERS :

The Photometric Voltages should be within 2V to 10V

Filters (nm)	Voltages (v)
340	5.706
405	5.973
505	6.203
545	6.409
580	6.457
630	6.654

Pipetting Check

Requirement : The C.V. should be less than 2% for Medium & Large, less than 3% for Small.

Self Test - Passed

Type	Target	Mean	SD	CV
SMALL	0.356	0.3945	0.0032	0.8083
MEDIUM	1.730	1.7000	0.0026	0.1500
LARGE	2.670	2.5721	0.0373	1.4500

Incubation Temperature

Range : 37°C + 0.5

On Board display the Temperature is 37.0 degree stable

On complete testing of the unit we found that the unit is working satisfactorily and used to carry out the necessary functions.

THIS IS AN ELECTRONICALLY GENERATED STATEMENT HENCE SIGNATURE NOT REQUIRED