

**Calibration**

**Calibration Monitor**

Status

RB History

RB Detail

Calibration History

Calibration Detail

Test Name

1.GLU

Type

Serum

Date/Time

04/01/2021 09:48

Passed

Reagent

Lot No.

Bottle No.

R1(R1-1)

2575

G334

R2(R2-1)

2575

G283

Cal Expiration Date

Reagent Blank

04/01/2021 09:45

Cal Type

AB

Measure Type

Rack


Formula

Y=AX+B

Factor

A = 4.1238E002

B = 0.0000E000



	Cal No.	CONC	OD
1	1	136	0.3298

Comment

Empty comment box.

Lot to Lot Calibration

RB/CAL Selection

Data Select

Comment

Graph Scale

Print

MEASURE 2

09/01/2021  
18:24

Calibration

Calibration Monitor

Status

RB History

RB Detail

Calibration History

Calibration Detail

Test Name

3.CREA

Type

Serum

Date/Time

09/01/2021 15:42

Passed

Reagent

Lot No.


Bottle No.

R1(R1-1)

2602

R2(R2-1)

2602



Cal Expiration Date

Reagent Blank 09/01/2021 15:40

Cal Type

AB

Measure Type

Rack

Formula

Y=AX+B

Factor

A = 8.0354E001

B = 0.0000E000

	Cal No.	CONC	OD
1	1	2.27	0.0283

Comment

Lot to Lot  
Calibration

RB/CAL  
Selection

Data Select

Comment

Graph Scale

Print



Calibration


Calibration Monitor

Status RB History RB Detail Calibration History Calibration Detail

Test Name: 19.ALT Type: Serum

Date/Time: 28/12/2020 09:06 Passed

Reagent	Lot No.	Bottle No.
R1(R1-1)	2553	8531
R2(R2-1)	2553	9747



Cal Expiration Date: \_\_\_\_\_  
 Reagent Blank: 28/12/2020 09:03  
 Cal Type: AB  
 Measure Type: Rack  
 Formula: Y=AX+B  
 Factor:  
 A = 4.4759E003  
 B = 0.0000E000

Cal No.	CONC	OD
1	85	0.0190

Comment

Lot to Lot Calibration RB/CAL Selection Data Select Comment Graph Scale Print

Calibration

Calibration Monitor

Status

RB History

RB Detail

Calibration History

Calibration Detail


Test Name: 18.AST

Date/Time: 24/12/2020 09:58

Passed

Type: Serum

Reagent Lot No. Bottle No.  
 R1(R1-1) 2530 D687  
 R2(R2-1) 2530 E645



Cal Expiration Date: \_\_\_\_\_  
 Reagent Blank: 24/12/2020 09:55  
 Cal Type: AB  
 Measure Type: Rack  
 Formula: Y=AX+B  
 Factor:  
 A = 4.6782E003  
 B = 0.0000E000

Cal No.	CONC	OD
1	108	0.0231

Comment: \_\_\_\_\_

**Calibration** **Calibration Monitor**

Status RB History RB Detail Calibration History Calibration Detail

Test Name: 15.TBILC Type: Serum

Date/Time: 25/12/2020 09:22 Passed

Reagent: Lot No.: Bottle No.:

R1(R1-1)

R2(R2-1)

Cal Expiration Date: \_\_\_\_\_


Reagent Blank: 25/12/2020 09:19

Cal Type: AB

Measure Type: Rack

Formula: Y=AX+B

Factor:  
 A = 3.5179E001  
 B = 0.0000E000



Cal No.	CONC	OD
1	3.40	0.0966


Comment

**Calibration** **Calibration Monitor**

Status RB History RB Detail Calibration History Calibration Detail

Test Name: 17.DBILC Type: Serum  
 Date/Time: 28/12/2020 09:06 Passed

Reagent Lot No. Bottle No.  
 R1(R1-1) 2574 B124  
 R2(R2-1)



Cal Expiration Date  
 Reagent Blank: 28/12/2020 09:03  
 Cal Type: AB  
 Measure Type: Rack  
 Formula: Y=AX+B  
 Factor:  
 A = 4.9914E001  
 B = 0.0000E000

Cal No.	CONC	OD
1	2.94	0.0589

Comment

Lot to Lot Calibration   
  RB/CAL Selection   
  Data Select   
  Comment   
  Graph Scale   
  Print

MEASURE 2

09/01/2021

18:27

Calibration

Calibration Monitor

Status

RB History

RB Detail

Calibration History

Calibration Detail

Test Name

20.ALP

Type

Serum

Date/Time

01/01/2021 09:40

Passed

Reagent

Lot No.

Bottle No.

R1(R1-1)

2547

F462

R2(R2-1)

2547

F786

Cal Expiration Date

Reagent Blank

01/01/2021 09:38

Cal Type

AB

Measure Type

Rack


Formula

Y=AX+B

Factor

A = 5.5871E003

B = 0.0000E000



Cal No.	CONC	OD
1	183	0.0328

Comment

Lot to Lot Calibration

RB/CAL Selection

Data Select

Comment

Graph Scale

Print



Calibration


Calibration Monitor

Status RB History RB Detail Calibration History Calibration Detail

Test Name: 8.TRIG Type: Serum

Date/Time: 06/01/2021 16:03 Passed

Reagent	Lot No.	Bottle No.
R1(R1-1)	2557	8664
R2(R2-1)	2557	8742



Cal Expiration Date: \_\_\_\_\_  
 Reagent Blank: 06/01/2021 16:01  
 Cal Type: AB  
 Measure Type: Rack  
 Formula: Y=AX+B  
 Factor:  
 A = 1.1852E003  
 B = 0.0000E000

Cal No.	CONC	OD
1	227	0.1915

Comment: \_\_\_\_\_



**Calibration**

**Calibration Monitor**

Status

RB History

RB Detail

Calibration History


Calibration Detail

Test Name: 9.CHOL

Type: Serum

Date/Time: 28/12/2020 09:05 Passed

Reagent: Lot No. 2561, Bottle No. 6528



Cal Expiration Date: \_\_\_\_\_  
 Reagent Blank: 28/12/2020 09:03  
 Cal Type: AB  
 Measure Type: Rack  
 Formula: Y=AX+B  
 Factor:  
 A = 9.2714E002  
 B = 0.0000E000

	Cal No.	CONC	OD
1	1	199	0.2146

Comment



MEASURE 2

09/01/2021

18:25

Calibration

Calibration Monitor

Status

RB History

RB Detail

Calibration History

Calibration Detail


Test Name 10.HDL

Type Serum

Date/Time 14/12/2020 09:25

Passed

Reagent	Lot No.	Bottle No.
R1(R1-1)	2436	8865
R2(R2-1)	2436	8294



Cal Expiration Date

Reagent Blank 14/12/2020 09:22

Cal Type AB

Measure Type Rack

Formula Y=AX+B

Factor

A = 3.0267E002

B = 0.0000E000

	Cal No.	CONC	OD
1	2	52	0.1718

Comment

Text area for comments

Lot to Lot Calibration

RB/CAL Selection

Data Select

Comment

Graph Scale

Print



Calibration

Calibration Monitor

Status

RB History

RB Detail

Calibration History


Calibration Detail

Test Name

Type Serum

Date/Time   Passed

Reagent	Lot No.	Bottle No.
R1(R1-1)	2555	C969
R2(R2-1)	2555	C825



Cal Expiration Date

Reagent Blank

Cal Type

Measure Type

Formula

Factor  
 A = 6.4893E001  
 B = 0.0000E000

	Cal No.	CONC	OD
1	1	7.5	0.1156

Comment

Calibration

Calibration Monitor

Status

RB History

RB Detail

Calibration History

Calibration Detail

Test Name 2.UREA

Type Serum


Date/Time 26/12/2020 09:55

Passed

Reagent Lot No. Bottle No.

R1(R1-1)

R2(R2-1)



Cal Expiration Date

Reagent Blank 26/12/2020 09:54

Cal Type AB

Measure Type Rack

Formula  $Y=AX+B$

Factor

A = 7.0374E002

B = 0.0000E000

Cal No.	CONC	OD
1	94.2	0.1339

Comment

Lot to Lot Calibration

RB/CAL Selection

Data Select

Comment

Graph Scale

Print



**Calibration**


**Calibration Monitor**

Status      RB History      RB Detail      Calibration History      Calibration Detail

Test Name: 5.CA  
 Date/Time: 03/01/2021 08:29  
 Reagent: R1(R1-1), R2(R2-1)  
 Lot No., Bottle No.

Type: Serum

Passed



Cal Expiration Date:  
 Reagent Blank: 03/01/2021 08:27  
 Cal Type: AB  
 Measure Type: Rack  
 Formula: Y=AX+B  
 Factor:  
 A = 1.4033E001  
 B = 0.0000E000

Cal No.	CONC	OD
1	9.9	0.7055

Comment

Calibration

Calibration Monitor

Status

RB History

RB Detail

Calibration History

Calibration Detail

Test Name

13.ALB

Type

Serum

Date/Time

08/01/2021 09:13

Passed


Reagent

Lot No.

Bottle No.

R1(R1-1)

R2(R2-1)



Cal Expiration Date

Reagent Blank 08/01/2021 09:11

Cal Type

AB

Measure Type

Rack

Formula

Y=AX+B

Factor

A = 6.5540E000

B = 0.0000E000

	Cal No.	CONC	OD
1	1	4.0	0.6103

Comment

Lot to Lot  
Calibration

RB/CAL  
Selection

Data Select

Comment

Graph Scale

Print



Calibration

Calibration Monitor

Status

RB History

RB Detail

Calibration History

Calibration Detail


Test Name 12.TP

Type Serum

Date/Time 08/01/2021 09:13

Passed

Reagent	Lot No.	Bottle No.
R1(R1-1)	2578	
R2(R2-1)	2578	



Cal Expiration Date

Reagent Blank 08/01/2021 09:11

Cal Type AB

Measure Type Rack

Formula Y=AX+B

Factor

A = 1.4579E001

B = 0.0000E000

	Cal No.	CONC	OD
1	1	6.8	0.4664

Comment

Lot to Lot  
Calibration

RB/CAL  
Selection

Data Select

Comment

Graph Scale

Print



# CALIBRATION CERTIFICATE

**INSTRUMENT NAME** : *Fully Automated Biochemistry Analyzer*

**MANUFACTURE NAME:** Beckman Coulter

**INSTRUMENT TYPE** : AU480

**INSTRUMENT SERIAL NUMBER** : 2016070383

**DATE OF CALIBRATION:** 5<sup>th</sup> March 2020

**NEXT DUE CALIBRATION ON:** 4<sup>th</sup> March 2021



Date : 5<sup>th</sup> March 2020  
 Customer: Satav Pathology Lab,Pune.  
 AU480 Calibration report  
 Serial No: 2016070383

Department: **BIOCHEMISTRY**  
 Instrument calibration checklist

Unit	Inspection Item	Contents	Functionality Test
AA Overall Body	Air filter	Washed	Passed
	D.I. water filter	Washed	Passed
	Water tank	Washed	Passed
	wash water metal filter	Washed	Passed
	Vaccum pump diaphragm	Checked	Passed
	Detergent peristaltic tube	Replaced	Passed
	wash station inlet water tubing	Washed	Passed
	wash station drain tube	Washed	Passed
	Power supply & other parts & their voltages	Cleaned & Measured	Passed (voltages are within range)
CA,CB,CC,CD,C E,CH, Racktransfer	Barcode reader	Cleaned & Checked	Passed
	general cleaning of all sensors	Cleaned & Checked	Passed
	CA Belt tension	Checked	Passed
	CA stop position	Checked	Passed
	Rack feeder rail (CA)	Checked	Passed
	Total rack transfer check	Checked	Passed
DA & DC	Reagent & stat compartment temperature measurement	Cleaned & Measured	Passed (within 8to 12 Deg)
	Barcode reader window	Cleaned & Checked	Passed
FA & FD	Whole module	Checked	Passed
Sample & reagent probe	Belt tension	Checked	Passed
	FA & FD wash plot	Cleaned	Passed
	All stop position alignment	Checked	Passed
FB	All nozzles	Cleaned & Checked	Passed
Cuvette wash station	FB nozzel alignment	Checked	Passed
	Function	Checked & Primed	Passed
FC	Mixer bar	Cleaned & Checked	Passed
Mixer station alignment	Mixer wash plot Mixer station alignment	Cleaned & Checked	Passed
GA	All cuvettes	Checked	Passed
Incubator & Cuvette wheel	Lamp	Checked	Passed
	Coolant level	Checked	Passed
	Photocal	Checked	Passed
SA	Syringe	Performed	Passed
DPR	Computer data backup	Checked	Passed
		Cleaned	Passed

1. All the mechanical adjustment & alignment are check & confirm using LED checker JIG provider by Beckman Coulter & the procedure as for the service manual.
2. All voltage are measure using an accurate digital multimeter.
3. The refrigeration & incubation temperature are measure using temperature control JIG provider by Beckman Coulter.
4. The functionality of the entire barcode, infrared, laser, reflective, magnetic, mechanical & optical sensors is checked & confirmed through the analyzer diagnostic menu provided in the software of the Beckman Coulter AU400.
5. All the pneumatic tubing's, & wash wells where body fluids are dispensed & passed is cleaned & washed with dilute sodium hypo chloride & deionised water.
6. All the water pneumatic tubing's are thoroughly washed & cleaned with dilute detergent & deionise water.
7. All the waste tubing's are thoroughly washed with concentrated sodium hypo chloride & deionised water.
8. The PH of the deionised water being used in the analyzer is measured & is reported to be NORMAL.

Inspection Engineer:-

Prashant Joshi

Prashant Joshi  
Sr.Customer Support Engineer