

Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROSCOPE

Certificate No:	20113000066	Issued On:	25/11/2020
Job Identification No:	ICS/C/11-3/66	Calibration Date:	21/11/2020
Ref. No:	SRF, Dated - 17/11/2020	Next Calibration Date:	20/11/2021

CALIBRATED FOR:	M/s Polo Lab, IVY Hospital, Opp. ST. Joseph School, Chandigarh Road Hoshiyarpur.
------------------------	---

EQUIPMENT DESCRIPTION

Name	Microscope	Visual Inspection	OK
Make/Model	Olympus	I.D. No.	IVY/HSPR/PATH/MS/02
Serial No.	---	Least Count	---
Range	4X to 100X	Calibration Site	In-Lab
Location	---		

ENVIRONMENTAL CONDITIONS	Temperature	20.0 °C ± 2.0° C	Humidity	50 ± 10 % RH
---------------------------------	--------------------	------------------	-----------------	--------------

STANDARD REFERENCE DETAILS

Traceable to National Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Slip Gauges-Set	Aditya/2938	TII/SGS/061/20	Tanson Instrument, Faridabad	24/06/2020	23/06/2022
Glass Scale	ICS/STD/91	TII/SGS/010/20	Tanson, Fbd	30/09/2020	29/09/2022

USED EQUIPMENT DETAILS	N/A
-------------------------------	-----

PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-D31
--	---

Parameter	Specified Magnification	STD Value	Measured Value On Magnification (mm) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2
Magnification	4 X	1 mm	4.0	± 0.18% of magnification
Magnification	10 X	1 mm	10.0	± 0.18% of magnification
Magnification	100 X	1 mm	99.5	± 0.18% of magnification

REMARKS:

The recommended date for next calibration is mentioned, as asked by the customer.

*** END OF REPORT ***

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 - Total liability of our Organisation is limited to the invoiced amount.
 - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

Authorised Signatory

RAJESH KAPOOR
Operation-Head