



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/GHPI/TD/D01	Date of Issue	: 19/12/2020
		Calibration Done On	: 19/12/2020
		Calibration Due On	: 18/12/2021

Customer Name	GENERAL HOSPITAL, PALAI.
---------------	---------------------------------

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Temperature Display	BABA	-	MES/TD/21/09	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/I02669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.


Environmental Conditions	Temperature	27 °C	Relative Humidity	52 %
--------------------------	-------------	-------	-------------------	------

PERFORMANCE DATA


Display	Measured	Tolerance
2.3 °C	2.2 °C	± 0.3 °C
4.2 °C	4.3 °C	± 0.3 °C
6.4 °C	6.3 °C	± 0.3 °C
8.1 °C	8.3 °C	± 0.3 °C

Calibration Status	Pass	✓	Fail
--------------------	------	---	------

Calibrated By


 Mahesh MC (Service Engineer)

Approved By


 Binoj Thomas (Quality Manager)