



## CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/GTSHKTA/REF/D01	Date of Issue	14/01/2021
		Calibration Done On	13/01/2021
		Calibration Due On	12/01/2022
Customer Name	Govt. TRIBAL SPECIALITY HOSPITAL, KOTTATHARA, PALAKKAD.		

### DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	LG	-	67095198	ICTC Lab

### DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/102669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

<b>Environmental Conditions</b>	Temperature	26 °C	Relative Humidity	52 %
---------------------------------	-------------	-------	-------------------	------

### PERFORMANCE DATA

Display	Measured	Tolerance
1.3 °C	1.4 °C	± 0.3 °C
2.4 °C	2.2 °C	± 0.3 °C
4.2 °C	4.3 °C	± 0.3 °C
7.2 °C	7.2 °C	± 0.3 °C

<b>Calibration Status</b>	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

Prabin KP (Service Engineer)

Approved By

Binoj Thomas (Quality Manager)

NB: Results reported are valid at the time of and under the stated conditions of measurements. This report shall not be reproduced without written approval by Medical Engineering & Services