Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No: 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone: 0452 - 2561691 Mobile: +91 94434 74610



GSTIN. 33AAHCA8053H1ZR

DL No. MDU/4205/20B,MDU/3971/21B

CIN: U33110TN2009PTC072140



CALIBRATION CERTIFICATE

Customer	M/S. Dharshini Clinical Lab & ECG,	Certificate No:	AMS/LAB/2889022
Name &	Moondrumavadi,	Issuing Date:	14/01/2021
Address:	Madurai.	Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	DCL/MMVD/MP/001
Make:	Lab	Department:	Laboratory
Model:	Variable	Calibrated On:	13/01/2021
Sl. No:	8097811	Calibration	* * 7
Type:	Variable	Due Suggested:	12/01/2022
Range &	5to50μl	DUC Condition	
Least Count:		on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	LM-2008863	08/08/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.	DUC Reading	Standard Reading (µI)		Correction
No.	(μl)	Trial 1	Trial 2	(μl)
1	5	5	5	0
2	. 10	10	10	0
3	20	20	20	0
4	50	50	50	0

Allowable Tolerance: ±1%

Note:

- 1. Calibration Points were selected as per customer request.
- 2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
- Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B.89.

aceagm@gmail.com



@ww.acemedical.in



Approved by: (Technical Manager)

B. 89:

93455 56140, 93455 56148