

Date: 14/02/2020

## CALIBRATION CERTIFICATE

This is to state that the hematology cell counter model **MICROS ES 60** bearing **Serial no: 904ESOH14853** at **CARE DIAGNOSTICS, BANGALORE, Karnataka state**, was calibrated with Calibrator on **14/02/2020**.

The Calibration Coefficient is automatically generated after calibration. Subsequently controls were processed and found them in respective range.

Calibrator Used: **ABX MINOCAL**  
Lot No.: **CX 443**  
Expiry Date: **05/03/2020**

Sr. No	Tested Parameter	Remarks
1	Startup / Blank Cycle	Passed
2	Repeatability/ Precision	Passed
3	Calibration	Passed
4	Quality Control	Passed

Next calibration cycle is due on **13/02/2021**.

YESHWANTH PADASHETTY

  
ENGINEER CUSTOMER SUPPORT  
Horiba Medical.