


# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Sri Vinayaga Micro Lab, No. 320, New Dharapuram Road, Palani – 624 601. Ph. 98656 39890	Certificate No:	AMS/LAB/2888992
		Issuing Date:	10/01/2021
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	SVM/LAB/CEN/001
Make:	Electro	Department:	Laboratory
Model:	Junior Lab Centrifuge	Calibrated On:	09/01/2021
Sl. No:	2015062311	Calibration Due Suggested:	08/01/2022
Type:	Manual	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 3500 / 0.65A.		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Lutron / DT-2234C	FSMS/CL/05/18-08/071-01	29/07/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

SL.No.	DUC Reading (Position)	Standard Reading (RPM)	Correction (RPM)
1	Position1	1253	0
2	Position2	1589	0
3	Position3	2456	0
4	Position4	2873	0
5	Position5	3297	0

Allowable Tolerance: Set by Customer


Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

*B. Olay*



Approved by:

(Technical Manager)

*B. Olay*





## CERTIFICATE OF CALIBRATION

Calibration Date : 19-12-2020

Next Calibration: 18-12-2021

Certificate No 2020110034

Name of the Instrument : Electra Junior Lab Centrifuge

Model : 6 Tubes

Running Speed Level : 3356 rpm

Max Speed Level : 4000 rpm

Make : Zenith lab Co. Ltd

Serial No 2015062311

Customer Address : Sri Vinayaka  
microlab Palani.

Calibrated by:

Sounderarajan



Reviewed by:

Charles

Declaration: We Declare that this centrifuge is calibrated and working properly.

NOTE: Only for Audit purpose.