



**CALIBRATION CERTIFICATE**

Certificate No.	MES/20-21/GDHNR/CEN/D02	Date of Issue :	31/12/2020
		Calibration Done On :	29/12/2020
		Calibration Due On :	28/12/2021

Customer Name	GOVT. DISTRICT HOSPITAL, NILAMBUR, MALAPPURAM.
---------------	------------------------------------------------

**DETAILS OF DEVICE UNDER CALIBRATION**

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Rotek	-	70100190	ICTC Laboratory

**DETAILS OF MASTER EQUIPMENT USED**

SI No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/102671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.


<b>Environmental Conditions</b>	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

**PERFORMANCE DATA**

<b>Result</b>	Obtained RPM
	3829 rpm

<b>Calibration Status</b>	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

  
 Vivek Mahendran (Service Engineer)

Approved By

  
 Binoj Thomas (Quality Manager)

