



### CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/GDHNR/REF/D11	Date of Issue : 31/12/2020
		Calibration Done On : 29/12/2020
		Calibration Due On : 28/12/2021

Customer Name	GOVT. DISTRICT HOSPITAL, NILAMBUR, MALAPPURAM.
---------------	--

#### DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	Whirlpool / Genius	-	70100980	ICTC

#### DETAILS OF MASTER EQUIPMENT USED

SI No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/102669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

<b>Environmental Conditions</b>	Temperature	27.5 °C	Relative Humidity	52 %
---------------------------------	-------------	---------	-------------------	------

#### PERFORMANCE DATA

Display	Measured	Tolerance
1.2 °C	1.4 °C	± 0.3 °C
2.3 °C	2.2 °C	± 0.3 °C
4.1 °C	4.2 °C	± 0.3 °C
7.2 °C	7.1 °C	± 0.3 °C

<b>Calibration Status</b>	Pass	✓	Fail
---------------------------	------	---	------

Calibrated By

Approved By

  
 Prabin K P (Service Engineer)

  
 Binoj Thomas (Quality Manager)

NB: Results reported are valid at the time of and under the stated conditions of measurements. This report shall not be reproduced without written approval by Medical Engineering & Services.