



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/GDHNR/CEN/D04	Date of Issue : 31/12/2020 Calibration Done On : 29/12/2020 Calibration Due On : 28/12/2021
-----------------	-------------------------	---

Customer Name	GOVT. DISTRICT HOSPITAL, NILAMBUR, MALAPPURAM.
---------------	--

DETAILS OF DEVICE UNDER CALIBRATION				
Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Remi / R-8C	GDLC- 8391	70100189	ICTC Laboratory

DETAILS OF MASTER EQUIPMENT USED					
Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/102671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021


The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
--------------------------	-------------	------	-------------------	------

PERFORMANCE DATA			
Set Reading	Display Reading	Measured Reading	Tolerance
Minimum	1049 rpm	1050 rpm	± 5 rpm
Medium	1125 rpm	1125 rpm	± 5 rpm
Maximum	1961 rpm	1962 rpm	± 5 rpm

Calibration Status	Pass	✓	Fail	
--------------------	------	---	------	--

Calibrated By


 Vivek Mahendran (Service Engineer)

Approved By


 Binoj Thomas (Quality Manager)



NB: Results reported are valid at the time of and under the stated conditions of measurements. This report shall not be reproduced without written approval by Medical Engineering & Services.