



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/WCHAA/CEN/D04	Date of Issue	: 20/08/2020
		Calibration Done On	: 14/08/2020
		Calibration Due On	: 13/08/2021

Customer Name	WOMEN AND CHILDREN HOSPITAL, ALAPPUZHA
---------------	---

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Remi / R-8c DX	-	80040085	Sneha Ward

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/MC/19-20/I02155-01	25/09/2020
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
--------------------------	-------------	------	-------------------	------


PERFORMANCE DATA


Set Reading	Display Reading	Measured Reading	Tolerance
Minimum	1640 rpm	1640 rpm	± 5 rpm
Medium	3300 rpm	3301 rpm	± 5 rpm
Maximum	4570 rpm	4571 rpm	± 5 rpm

Calibration Status	Pass	✓	Fail
--------------------	------	---	------

Calibrated By

Approved By


 Ankit Vinay (Service Engineer)


 Binoj Thomas (Quality Manager)