



### CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/WCHAA/REF/D11	Date of Issue	: 20/08/2020
		Calibration Done On	: 14/08/2020
		Calibration Due On	: 13/08/2021

Customer Name	<b>WOMEN AND CHILDREN HOSPITAL, ALAPPUZHA</b>
---------------	---

#### DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	Samsung	-	68040840	Sneha Ward

#### DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/TH/19-20/I02155-01	25/09/2020
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.


<b>Environmental Conditions</b>	Temperature	27 °C	Relative Humidity	55 %
---------------------------------	-------------	-------	-------------------	------

#### PERFORMANCE DATA


Display	Measured	Tolerance
1.3 °C	1.3 °C	± 0.3 °C
2.4 °C	2.3 °C	± 0.3 °C
4.2 °C	4.2 °C	± 0.3 °C
7.2 °C	7.2 °C	± 0.3 °C

<b>Calibration Status</b>	Pass	✓	Fail
---------------------------	------	---	------

Calibrated By

  
 Amal Sasi (Service Engineer)

Approved By

  
 Binoj Thomas (Quality Manager)

POCO