

## LAB MONTHLY SUMMARY



Lab Name **GREEN LIFE DIAGNOSTIC CARE** Lab No **6158**  
 Month **September** Year **2020**  
 Constituent Group **Chemistry II**



Date of Result Entered : 20/09/2020

Date of Report Published : 07/10/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Robonik ( Semi automatic / Fully Automatic )	375	221.88	10.22	22.68	234.5 mg/dl	0.56	2.34
2	UREA I	Urease UV / GLDH	Robonik ( Semi automatic / Fully Automatic )	94	125.25	14.41	18.05	125.9 mg/dl	0.04	3.72
3	CREATININE I	Jaffes Kinetic - Alkaline picrate	Robonik ( Semi automatic / Fully Automatic )	281	1.41	11.88	0.17	1.4 mg/dl	-0.06	0.02
4	T.BILIRUBIN I	Diazonium salt ( Colorimetric ) / Jendrassik	Robonik ( Semi automatic / Fully Automatic )	157	5.11	15.10	0.77	5.4 mg/dl	0.38	0.12
5	T-PROTEIN I	Biuret - Colorimetric	Robonik ( Semi automatic / Fully Automatic )	158	5.44	10.17	0.55	5.6 g/dl	0.29	0.09
6	ALBUMIN I	BCG - colorimetric	Agape	104	3.24	8.40	0.27	3.5 g/dl	0.96	0.05
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Robonik ( Semi automatic / Fully Automatic )	188	7.23	14.78	1.07	11.1 mg/dl	3.62	0.16
8	CHOLESTEROL I	CHOD-PAP	Robonik ( Semi automatic / Fully Automatic )	318	109.16	10.70	11.68	65 mg/dl	-3.78	1.31
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Robonik ( Semi automatic / Fully Automatic )	166	100.74	17.54	17.67	110 mg/dl	0.52	2.74

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*