

VIEW LAB MONTHLY SUMMARY

Lab Name: SIRE DIAGNOSTIC LAB
Lab No: 13577
Month: December
Year: 2020
Consultant Group: Chemistry II

[Details About Richard Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SIZ](#)

[Click on the analyte to view Graphical Data](#)

[All Analytes Result](#) [Print](#)

Date of Result Entered: 18/12/2020

Date of Report Published: 31/12/2020

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	IQ			
1	GLUCOSE I	GOD-POD	Transasia / Erba	523	144.42	10.31	14.89	157.7 mg/dl	0.89	1.18
2	UREA I	Urease UV / GLDM	Transasia / Erba	524	81.07	13.19	10.69	86.4 mg/dl	0.50	0.88
3	CREATININE I	Enzymatic Colorimetric	Transasia / Erba	127	4.20	0.21	0.34	4.4 mg/dl	0.58	0.06
4	T.BILIRUBIN I	Diazonium salt (Colorimetric) / Jendrasik	Transasia / Erba	512	2.10	15.60	0.33	1.8 mg/dl	0.91	0.03
5	T.PROTEIN I	Buret - colorimetric	Transasia / Erba	463	5.04	9.93	0.50	4.9 g/dl	0.28	0.65
6	ALBUMIN I	BCG - colorimetric	Transasia / Erba	462	3.04	9.40	0.29	2.7 g/dl	-1.19	0.03
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Transasia / Erba	500	4.13	17.35	0.72	4 mg/dl	0.18	0.06
8	CHOLESTEROL I	CHOD-PAP	Transasia / Erba	547	98.47	11.41	11.23	107 mg/dl	0.76	0.96
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	508	96.95	19.98	19.37	79.3 mg/dl	-0.91	1.72

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent
Between +1.0 to +2.0	Good
Between +2.0 to +3.0	Accept with caution: Warning Signal.
Beyond +3.0	Unacceptable performance: Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out