

CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/01068

Page: 1 of 2

Customer Name:	SREE METRO DIAGNOSTIC CENTRE	Customer Address:	16/41, KRISHNA ROAD, NEW PERUNGALATHUR, CHENNAI-63
Location:	LAB-1	Serial No:	17606607
Description:	MICROPIPETTE	Tag/Asset No:	SMDC/LAB-1/PIP/04
Make:	BIOHIT	Date of Calibration:	30-09-2020
Model:	PROLINE	Date of Due:	29-09-2021
Range:	10 µL		FIXED

Traceability Details:


S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021

Environmental Conditions and Standard Operating Procedure Details:

ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS			
Temperature	24°C	Calibration Done at	ONSITE
Humidity(40-70% RH)	55 %	Calibration Procedure No	--

REMARKS:

Test Report : Attached

Calibrated by: 
M.BALAJI
(Calibration Engineer)

Approved by: 
C.SHANMUGARAJ
(Technical Manager)



Result of Calibration: SBS/TR/20-21/01068


Page: 2 of 2

TEST REPORT				
Units	Nominal Values	Measured Values	Tolerance	Result
µl	10	10.0	±0.5 %	PASS



CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: 

Date: 30-09-2020