

## CALIBRATION REPORT

**Details of the Unit Under Calibration:**

**Calibration Certificate No: SBS/CL/20-21/01070**

Page: 1 of 2

<b>Customer Name:</b>	SREE METRO DIAGNOSTIC CENTRE	<b>Customer Address:</b>	16/41, KRISHNA ROAD, NEW PERUNGALATHUR, CHENNAI-63
<b>Location:</b>	LAB-1	<b>Serial No:</b>	17573869
<b>Description:</b>	MICROPIPETTE	<b>Tag/Asset No:</b>	SMDC/LAB-1/PIP/06
<b>Make:</b>	BIOHIT	<b>Date of Calibration:</b>	30-09-2020
<b>Model:</b>	PROLINE	<b>Date of Due:</b>	29-09-2021
<b>Range:</b>	500 µL		FIXED

**Traceability Details:**

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021

**Environmental Conditions and Standard Operating Procedure Details:**

ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS			
<b>Temperature</b>	24°C	<b>Calibration Done at</b>	ONSITE
<b>Humidity(40-70% RH)</b>	55 %	<b>Calibration Procedure No</b>	--

**REMARKS:**

Test Report : Attached

Calibrated by:

  
M.BALAJI  
(Calibration Engineer)

Approved by:

  
C.SHANMUGARAJ  
(Technical Manager)



Result of Calibration: SBS/TR/20-21/01070

Page: 2 of 2

TEST REPORT				
Units	Nominal Values	Measured Values	Tolerance	Result
$\mu\text{l}$	500	500.1	$\pm 0.5\%$	PASS



**CALIBRATION RESULT:**

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign:



Date: 30-09-2020