

CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/01066

Page: 1 of 2

Customer Name:	SREE METRO DIAGNOSTIC CENTRE	Customer Address:	16/41, KRISHNA ROAD, NEW PERUNGALATHUR, CHENNAI-63
Location:	LAB-1	Serial No:	1309624500
Description:	MICROPIPETTE	Tag/Asset No:	SMDC/LAB-1/PIP/02
Make:	FINNPIPETTE	Date of Calibration:	30-09-2020
Model:	NA	Date of Due:	29-09-2021
Range:	150-1050 µL		FIXED

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021


Environmental Conditions and Standard Operating Procedure Details:

ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS			
Temperature	24°C	Calibration Done at	ONSITE
Humidity(40-70% RH)	55 %	Calibration Procedure No	--

REMARKS:

Test Report : Attached

Calibrated by:


M. BALAJI
 (Calibration Engineer)

Approved by:



C. SHANMUGARAJ
 (Technical Manager)



TEST REPORT				
Units	Nominal Values	Measured Values	Tolerance	Result
μ l	150	149.9	$\pm 0.5\%$	PASS
μ l	200	199.9	$\pm 0.5\%$	PASS
μ l	500	500.1	$\pm 0.5\%$	PASS
μ l	1000	1000.1	$\pm 0.5\%$	PASS
μ l	1050	1050.1	$\pm 0.5\%$	PASS

CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: 

Date: 30-09-2020