

Lab Monthly Summary
 Lab Name IQURE BIOMEDICAL SYSTEMS PVT LTD
 Month December
 Constituent Group Chemistry II

Lab No 13278
 Year 2020

Date of Result Entered : 17/12/2020
 Date of Report Published : 31/12/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U																																																							
						CV	SD																																																										
1	GLUCOSE I	GOD-POD	Mindray	128	145.77	9.57	13.95	136.2 mg/dl	-0.69	2.47																																																							
2	UREA I	Urease UV / GLDH	Mindray	118	82.10	12.19	10.01	83.6 mg/dl	0.15	1.84																																																							
3	CREATININE I	OTHERS (Any Other Principles / Methods)	Any Analyser	124	3.94	12.91	0.51	3.8 mg/dl	-0.28	0.09																																																							
4	T.BILIRUBIN I	Diazonium salt (Colorimetric) / Jendrassik	Mindray	83	2.11	14.42	0.30	2.03 mg/dl	-0.26	0.07																																																							
5	T-PROTEIN I	Biuret - Colorimetric	Mindray	112	4.93	10.05	0.50	4.6 g/dl	-0.67	0.09																																																							
6	ALBUMIN I	BCG - colorimetric	Mindray	101	3.05	10.31	0.32	2.57 g/dl	-1.52	0.06																																																							
7	URIC ACID I	Enzymatic / Uricase Colorimetric	Mindray	91	4.03	19.71	0.80	3.18 mg/dl	-1.07	0.17																																																							
8	CHOLESTEROL I	CHOD-PAP	Mindray	115	99.66	9.71	9.68	100.3 mg/dl	0.07	1.81																																																							
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	109	110.73	12.59	13.94	115.6 mg/dl	0.35	2.67																																																							
<table border="1"> <thead> <tr> <th colspan="11">SDI Range Interpretation</th> </tr> </thead> <tbody> <tr> <td colspan="11">Within -1.0 to +1.0 Excellent.</td> </tr> <tr> <td colspan="11">Between ±1.0 to ±2.0 Good.</td> </tr> <tr> <td colspan="11">Between ±2.0 to ±3.0 Accept with caution. Warning Signal.</td> </tr> <tr> <td colspan="11">Beyond ±3.0 Unacceptable performance. Action Signal.</td> </tr> </tbody> </table>											SDI Range Interpretation											Within -1.0 to +1.0 Excellent.											Between ±1.0 to ±2.0 Good.											Between ±2.0 to ±3.0 Accept with caution. Warning Signal.											Beyond ±3.0 Unacceptable performance. Action Signal.										
SDI Range Interpretation																																																																	
Within -1.0 to +1.0 Excellent.																																																																	
Between ±1.0 to ±2.0 Good.																																																																	
Between ±2.0 to ±3.0 Accept with caution. Warning Signal.																																																																	
Beyond ±3.0 Unacceptable performance. Action Signal.																																																																	