

## Mindray BS-240 Auto Chemistry Analyzer – PQ Documents

Customer Address: **I Qure path labs (RJ nagar),** Contact person : **Dr. Manjunath**  
No. #13/2, 2nd floor, KHM block, Telephone no : 09945568004  
1st main road, RJ nagar,  
Bangalore .  
Karnataka.

Serial no : **YX- 94002462**

Performed on : 13/07/19

### Performance Qualification

#### CALIBRATION:

The Analyzer has been calibrated with Mindray Multi sera calibrator for D.billirubin , ALT, T.billirubin, Creatinine-S,TC, Glucose-G, Calcium, TG, ALP, AST, Total protein, Urea.

Calibrator Manufacturer : Mindray  
Calibrator name : Multi Sera calibrator  
Lot Numer : 150119001  
Expiry date : 20/02/2021

#### QUALITY CONTROL CHECK - NORMAL CONTROL

Control manufacturer : Mindray  
QC Name : Multi Control Sera N  
Lot Number : 050218010  
Expiry Date : 09/12/2020

Carried out Multi Control Sera N found values are within specified range.

#### REMARKS:

The system meets all criteria outlined in this performance qualification documents. All the required parameter has been checked and instrument is ready for regular use.

Performer : **Santhosh . R**  
Designation : Sr.Application specialist

Reviewer : **Muthu Vigneshwar . K.P**  
Designation : Application Manager

Date : 19/07/2019