

Sl.No	Parameter	Reference Value	NABL Lab (Metropolis)	Deviation Observed	Remarks
1	Physical Examination				
2	Volume	10	5	Not Significant	ILC Found Satisfactory
3	Colour	Pale Yellow	Pale Yellow	Not Significant	ILC Found Satisfactory
4	Appearance	Clear	Clear	Not Significant	ILC Found Satisfactory
5	Specific Gravity	1.03	1.03	Not Significant	ILC Found Satisfactory
6	pH	6	5.5	Not Significant	ILC Found Satisfactory
7	Chemical Examination				
8	Urine Albumin	Nil	Nil	Not Significant	ILC Found Satisfactory
9	Urine Sugar	Nil	Nil	Not Significant	ILC Found Satisfactory
10	Urine Nitrates	Negative	Negative	Not Significant	ILC Found Satisfactory
11	Microscopic Examination				
12	RBC's	Absent	0-1	Not Significant	ILC Found Satisfactory
13	Leucocytes	02-Apr	4	Not Significant	ILC Found Satisfactory
14	Epithelial Cells	01-Feb	3.4	Not Significant	ILC Found Satisfactory