

LAB MONTHLY SUMMARY



Lab Name **SRI RAMANA LAB X-RAY & ECG.** Lab No **2957**
 Month **December** Year **2020**
 Constituent Group **Chemistry II**



Date of Result Entered : 16/12/2020

Date of Report Published : 31/12/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	CPC (Turbochem 100 / 240 Prime	28	143.07	10.20	14.59	149 mg/dl	0.41	5.52
2	UREA I	Urease UV / GLDH	CPC (Turbochem 100 / 240 Prime	27	84.71	11.42	9.68	88.97 mg/dl	0.44	3.72
3	CREATININE I	Jaffes End point	Any Analyser	361	3.94	13.67	0.54	3.62 mg/dl	-0.59	0.06
4	T.BILIRUBIN I	Diazonium salt (Colorimetric) / Jendrassik	CPC (Turbochem 100 / 240 Prime	22	2.03	7.72	0.16	1.76 mg/dl	-1.72	0.07
5	T-PROTEIN I	Biuret - Colorimetric	CPC (Turbochem 100 / 240 Prime	23	4.76	18.26	0.87	3 g/dl	-2.02	0.36
6	ALBUMIN I	BCG - colorimetric	CPC (Turbochem 100 / 240 Prime	24	2.86	6.99	0.20	2.83 g/dl	-0.15	0.08
7	URIC ACID I	Enzymatic / Uricase Colorimetric	CPC (Turbochem 100 / 240 Prime	23	4.03	11.85	0.48	3.9 mg/dl	-0.27	0.20
8	CHOLESTEROL I	CHOD-PAP	CPC (Turbochem 100 / 240 Prime	28	93.81	8.40	7.88	90 mg/dl	-0.48	2.98
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	CPC (Turbochem 100 / 240 Prime	30	99.63	15.94	15.88	72 mg/dl	-1.74	5.80

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****