

Calibration Certificate

Following voltages have been checked & adjusted on Blood cell counter Swelab Alfa Plus (Sr. No. 101651) installed at Star Health Diagnostic & Imaging Centre, G2/82, Rohini sector -16, Delhi – 110089.

- Power Supply : 24V DC Voltages are OK.
- Checked the Gain and Offset of Analog Board and found OK.
- Checked the background values of different parameters found OK.
- All instrument thru' the Service Program and found OK.
- Checked the Pumps & Valves and found OK.
- Calibrate the instrument and Run control. Control results were OK
- Next Calibration will be performed on **31/12/2021**.



For AAMX Biomedicals,
Nitin Choudhary
Delhi

Date : **01/01/2021**

: : : :
:
: