



ARKRAY HEALTHCARE PVT. LTD.
 Plot No. 336, 338, 340, Karamba, S.G. Road,
 Scheme 94/291, Surat, India. Telephone: 0261-2397719 to 14
 Facsimile: 0261-2397719, URL: <http://www.arkray.com>

PERFORMANCE QUALIFICATION (PO)

Serial Number	1250619163	Date: 18/09/2020
---------------	------------	------------------

Calibration: Calibration done with calibrator C.f.a.s.

Verification of the instrument results is done and in range.

Precicontrol Clinichem Multi 2 Lot (250280)	Results	Range of Control
Glucose	233.00	(195-267) mg/dl
Cholesterol	173.83	(152-206) mg/dl
SGOT	143.80	(113-161)U/L
Billirubit Total	3.92	(3.12-4.50)mg/dl
Urea	114.00	(99-135)mg/dl
ALP	206.43	(168-240)U/L

The results obtained are satisfactory. This is to certify that AutoChem Xact Pro is working OK & results are within recommended range

 For, Arkray Healthcare Pvt Ltd.		Customer's Signature and Stamp
------------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------	--------------------------------



arkray healthcare Pvt. Ltd.
Pka no. 116, 131, 141, Road no. 1, G.I.D.C.,
Sector 194 230, (Surat), India. Telephone: 0261-239 7712 to 15
Facsimile: 0261-239 7719, URL: <http://www.arkray.com>

Calibration Certificate

Customer Name	STAR HEALTH DIAGNOSTIC & IMAGING CENTRE
Instrument Name	AUTO CHEM XACT PRO
Installation Date	18-09-2020
Serial Number	1250619163
Calibrated On	18-09-2020
Next Calibration Due	17-09-2021

Filter Check

Filter (nm)	Status	Filter (nm)	Status
340 nm	OK	405 nm	OK
450 nm	OK	510 nm	OK
546 nm	OK	578 nm	OK
620 nm	OK	660 nm	OK
690 nm	OK		

On basis of above mention checks, the instrument is functioning satisfactory.
If any major breakdown occur instrument to be re-validated.

Best Regards,

Atul Kumar Tiwari(SE)
Arkray Healthcare Pvt Ltd