



Cert. No : CC-2788



#40, 2nd Floor, 2nd Street,
Padmavathi Nagar, Chromepet,
Chennai, Tamil Nadu - 600 044.

CALIBRATION CERTIFICATE

MCAL/FOR/QD-12

Page # 1 of 1

PRF #	: MCAL/PRF/012166	Date of Calibration	: 21-Jan-21
Certificate #	: MSIR/410121/66-001	Recom. Due Date	: 20-Jan-22

DETAILS OF USER CUSTOMER

M/S. Diacare Diagnostics,
2/751, Veppanapalli,
Krishnagiri(D), Tamil Nadu.

ULR #	: ULR-CC27882100000828F
Calibrated at	: Onsite
Date of Receipt	: 21-Jan-21
Cond. On Receipt	: Satisfactory
Date of Issue	: 22-Jan-21

DETAILS OF TEST / UNIT UNDER CALIBRATION

Description	: Refrigerator	Identification No.	: 6
Range	: 2° to 8°C	Serial No.	: 170901123DR01315
Least Count	: --	Make	: Godrej
Cal./Working Range	: --	Model / Type	: RD Edge 185E2H
Location	: Biochemistry	Accuracy	: As per Manual

DETAILS OF MASTER INSTRUMENT USED



Description	Identification #	Valid upto	Certificate #
RTD	MCAL/INS/TH-07	14-May-21	MSIR/410520/INT-026
Digital Thermometer	MCAL/INS/ET-06	13-Sep-21	MSIR/410920/INT-055


MFC #	: MCAL/MFC/TH-07
Details of Env.	: Temperature: 25±5°C Relative Humidity: 40-70 % RH

CALIBRATION RESULT				
Sl.No.	TEST READING (°C)	STD READING (°C)	ERROR (°C)	UNCERTAINTY ±°C
1	2.0	2.1	-0.1	0.4
2	6.0	6.2	-0.2	0.4
3	8.0	8.3	-0.3	0.4

Remarks :

1. The reported Expanded Uncertainty is calculated at 95.45 % C.L with coverage factor k=2
2. The reading given was the average of three measurements.

Issued by :	 MCAL - Engineer Sangeetha.B	Authorised by :	 QM / TM Kamalanathan
-------------	---	-----------------	--



MSIR INDIA PRIVATE LIMITED

(The Madras Science and Industrial Resources)

Phone :044 2265 2634, +91 9789877567, Email: calibration@msirindia.com Web: www.msirindia.com